



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109830

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/01/2020	17:30
2	CAL2	1	Water	NA	NA	20.1	7.00	07/01/2020	17:31
3	CAL3	1	Water	NA	NA	20.1	10.03	07/01/2020	17:33
4	ICV	1	Water	NA	NA	20.1	7.02	07/01/2020	17:35
5	CCV1	1	Water	NA	NA	20.2	2.02	07/01/2020	17:36
6	L3188-04	1	Solid	20.03	20	21.9	9.13	07/01/2020	17:39
7	L3188-04DUP	1	Solid	20.04	20	21.8	9.14	07/01/2020	17:40
8	L3197-03	1	Solid	20.02	20	21.8	5.45	07/01/2020	17:42
9	L3197-09	1	Solid	20.03	20	21.7	5.72	07/01/2020	17:44
10	CCV2	1	Water	NA	NA	20.2	12.02	07/01/2020	17:45

VF5109830

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3197 WorkList ID : 140083 Department : Wet-Chemistry Date : 07-01-2020 13:19:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/01/2020	Solid	L3188-04	Corrosivity	Cool 4 deg C	PSEG05	P23	TR-04-070120	07/01/2020	9045D
06/30/2020	Solid	L3197-03	Corrosivity	Cool 4 deg C	PSEG01	P22	MH-1-COMP	06/30/2020	9045D
07/01/2020	Solid	L3197-09	Corrosivity	Cool 4 deg C	PSEG01	P22	MH-2-COMP	07/01/2020	9045D

Date/Time 07-01-2020 16:12:00
Received by: J. J. (wet-chem)
Relinquished by: KS (wet-chem)

Date/Time 07-01-2020 18:10:00
Received by: KS (wet-chem)
Relinquished by: J. J. (wet-chem)