



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109831

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/01/2020	17:30
2	CAL2	1	Water	NA	NA	20.1	7.00	07/01/2020	17:31
3	CAL3	1	Water	NA	NA	20.1	10.03	07/01/2020	17:32
4	ICV	1	Water	NA	NA	20.1	6.99	07/01/2020	17:35
5	CCV1	1	Water	NA	NA	20.1	2.01	07/01/2020	17:45
6	L3173-01	1	Solid	20.02	20	21.6	7.81	07/01/2020	17:46
7	L3173-01DUP	1	Solid	20.03	20	21.7	7.82	07/01/2020	17:48
8	CCV2	1	Water	NA	NA	20.1	12.02	07/01/2020	17:50

LB 109831

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3173 WorkList ID : 140096 Department : Wet-Chemistry Date : 07-01-2020 17:10:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/01/2020	Solid	L3173-01	pH	Cool 4 deg C	PSEG01	P22	13854	07/01/2020	9045D

Date/Time 07-01-2020 17:15
Received by: Jg WAT-CHM
Relinquished by: KS (wetchem)

Date/Time 07-01-2020 18:40
Received by: KS (wetchem)
Relinquished by: Jg WAT-CHM