



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109842

Slope : 97.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/02/2020	12:25
2	CAL2	1	Water	NA	NA	20.1	7.00	07/02/2020	12:26
3	CAL3	1	Water	NA	NA	20.1	10.03	07/02/2020	12:27
4	ICV	1	Water	NA	NA	20.1	7.01	07/02/2020	12:30
5	CCV1	1	Water	NA	NA	20.1	2.01	07/02/2020	12:32
6	L3209-03	1	Water	NA	NA	21.8	5.47	07/02/2020	12:35
7	L3209-03DUP	1	Water	NA	NA	21.9	5.48	07/02/2020	12:36
8	CCV2	1	Water	NA	NA	20.3	12.02	07/02/2020	12:39