



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109854

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/02/2020	18:05
2	CAL2	1	Water	NA	NA	20.2	7.00	07/02/2020	18:06
3	CAL3	1	Water	NA	NA	20.2	10.03	07/02/2020	18:07
4	ICV	1	Water	NA	NA	20.3	6.99	07/02/2020	18:09
5	CCV1	1	Water	NA	NA	20.3	2.02	07/02/2020	18:10
6	L3202-02	1	Solid	20.03	20	24.6	7.81	07/02/2020	18:11
7	L3204-02	1	Solid	20.02	20	23.1	6.09	07/02/2020	18:12
8	L3205-02	1	Solid	20.03	20	22.5	6.85	07/02/2020	18:15
9	L3205-04	1	Solid	20.04	20	22.6	6.81	07/02/2020	18:16
10	L3205-04DUP	1	Solid	20.03	20	22.7	6.82	07/02/2020	18:19
11	CCV2	1	Water	NA	NA	20.1	12.02	07/02/2020	18:20

LB104854

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3204 WorkList ID : 140130 Department : Wet-Chemistry Date : 07-02-2020 14:08:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/02/2020	Solid	L3202-02	Corrosivity	Cool 4 deg C	PSEG01	P32	RO-COMP	07/02/2020	9045D
07/02/2020	Solid	L3204-02	Corrosivity	Cool 4 deg C	PSEG01	P31	WF-COMP	07/02/2020	9045D
07/02/2020	Solid	L3205-02	Corrosivity	Cool 4 deg C	PSEG01	P31	12013	07/02/2020	9045D
07/02/2020	Solid	L3205-04	Corrosivity	Cool 4 deg C	PSEG01	P31	VNJ-214	07/02/2020	9045D

Date/Time 07/02/2020 16:20
 Received by: Jp (wet chem)
 Relinquished by: Jp (chem)

Date/Time 07/02/2020 18:00
 Received by: Jp (wet chem)
 Relinquished by: Jp (chem)