



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109856

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/02/2020	18:05
2	CAL2	1	Water	NA	NA	20.3	7.01	07/02/2020	18:06
3	CAL3	1	Water	NA	NA	20.3	10.03	07/02/2020	18:08
4	ICV	1	Water	NA	NA	20.3	6.99	07/02/2020	18:10
5	CCV1	1	Water	NA	NA	20.2	2.01	07/02/2020	18:25
6	L3206-05	1	Solid	20.03	20	24.1	8.83	07/02/2020	18:36
7	L3206-05DUP	1	Solid	20.04	20	24.2	8.84	07/02/2020	18:37
8	CCV2	1	Water	NA	NA	20.3	12.02	07/02/2020	18:38

JB 109856

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3206 S WorkList ID : 140134 Department : Wet-Chemistry Date : 07-02-2020 17:11:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/02/2020	Solid	L3206-05	pH	Cool 4 deg C	PSEG01	P33	31044	07/02/2020	9045D

Date/Time 07/02/2020 16:20
Received by: JB
Relinquished by: JB

Date/Time 07/02/2020 18:10
Received by: JB
Relinquished by: JB (n12m)