


LB109870

ISM02.4

INSTRUMENT ID: CY2

Linear



A= 0.0000e+000

B= 2.1784e-004

C= -3.1003e-002

Rho= 0.9999323

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.00	0.000	-0.027	-0.027	20	0.000	20					1
0.05	0.050					223					1
0.20	0.200	0.202	0.002	1070	0.0 %	1070					3
2.50	2.500	2.576	0.076	11969	0.0 %	11969					1
5.00	5.000	4.934	-0.066	22790	0.0 %	22790					0
7.50	7.500	7.507	0.007	34603	0.0 %	34603					0
10.0	10.000	10.008	0.008	46084	0.0 %	46084					0

LB109870
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	20	-		0 ISM02.4	PPB	7/6/2020 14:17	Std
	0.2 S0.2	1070	-		0.2 ISM02.4	PPB	7/6/2020 14:20	Std
	2.5 S2.5	11969	-		2.5 ISM02.4	PPB	7/6/2020 14:22	Std
	5 S5.0	22790	-		5 ISM02.4	PPB	7/6/2020 14:24	Std
	7.5 S7.5	34603	-		7.5 ISM02.4	PPB	7/6/2020 14:26	Std
	10 S10.0	46084	-		10 ISM02.4	PPB	7/6/2020 14:29	Std
ICV	ICV	18325	3.9609 -		ISM02.4	PPB	7/6/2020 14:31	SMPL
ICB	ICB	-72	-0.0467 -		ISM02.4	PPB	7/6/2020 14:34	SMPL
CCV043	CCV043	22271	4.8205 -		ISM02.4	PPB	7/6/2020 14:36	SMPL
CCB043	CCB043	-84	-0.0493 -		ISM02.4	PPB	7/6/2020 14:38	SMPL
PB129977BL	PBW977	-35	-0.0386 -		ISM02.4	PPB	7/6/2020 14:40	SMPL
L3092-01	MBFW32	171	0.0062 -		ISM02.4	PPB	7/6/2020 14:43	SMPL
L3092-02	MBFW37	-44	-0.0406 -		ISM02.4	PPB	7/6/2020 14:45	SMPL
L3092-03	MBFW38	-31	-0.0378 -		ISM02.4	PPB	7/6/2020 14:47	SMPL
L3092-04	MBFW39	-44	-0.0406 -		ISM02.4	PPB	7/6/2020 14:49	SMPL
L3092-05	MBFW40	-32	-0.038 -		ISM02.4	PPB	7/6/2020 14:52	SMPL
L3092-06	MBFW35	70	-0.0158 -		ISM02.4	PPB	7/6/2020 14:54	SMPL
L3092-07	MBFW35D	9	-0.029 -		ISM02.4	PPB	7/6/2020 14:56	SMPL
L3092-08	MBFW35S	4598	0.9706 -		ISM02.4	PPB	7/6/2020 14:59	SMPL
L3092-09	MBFW36	-9	-0.033 -		ISM02.4	PPB	7/6/2020 15:01	SMPL
L3092-10	MBFW42	-45	-0.0408 -		ISM02.4	PPB	7/6/2020 15:03	SMPL
L3092-11	MBFW44	-32	-0.038 -		ISM02.4	PPB	7/6/2020 15:05	SMPL
L3092-12	MBFW45	-33	-0.0382 -		ISM02.4	PPB	7/6/2020 15:08	SMPL
L3092-13	MBFW46	-33	-0.0382 -		ISM02.4	PPB	7/6/2020 15:10	SMPL
PB129978BL	PBW978	-10	-0.0332 -		ISM02.4	PPB	7/6/2020 15:12	SMPL
L3169-01	MBFW48	-24	-0.0362 -		ISM02.4	PPB	7/6/2020 15:14	SMPL
L3169-02	MBFW49	147	0.001 -		ISM02.4	PPB	7/6/2020 15:17	SMPL
L3169-03	MBFW50	79	-0.0138 -		ISM02.4	PPB	7/6/2020 15:19	SMPL
L3169-04	MBFW51	-51	-0.0421 -		ISM02.4	PPB	7/6/2020 15:21	SMPL
L3169-05	MBFW53	-34	-0.0384 -		ISM02.4	PPB	7/6/2020 15:23	SMPL
CCV044	CCV044	21455	4.6427 -		ISM02.4	PPB	7/6/2020 15:26	SMPL
CCB044	CCB044	-90	-0.0506 -		ISM02.4	PPB	7/6/2020 15:28	SMPL
L3169-06	MBFW54	11	-0.0286 -		ISM02.4	PPB	7/6/2020 15:30	SMPL
L3169-07	MBFW55	246	0.0226 -		ISM02.4	PPB	7/6/2020 15:33	SMPL
L3169-08	MBFW56	-64	-0.0449 -		ISM02.4	PPB	7/6/2020 15:35	SMPL
L3169-09	MBFW56D	-45	-0.0408 -		ISM02.4	PPB	7/6/2020 15:37	SMPL
L3169-10	MBFW56S	4518	0.9532 -		ISM02.4	PPB	7/6/2020 15:39	SMPL
CCV045	CCV045	21366	4.6233 -		ISM02.4	PPB	7/6/2020 15:42	SMPL
CCB045	CCB045	-87	-0.05 -		ISM02.4	PPB	7/6/2020 15:44	SMPL