



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109882

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	07/07/2020	08:10
2	CAL2	1	Water	NA	NA	20.4	7.00	07/07/2020	08:11
3	CAL3	1	Water	NA	NA	20.4	10.03	07/07/2020	08:12
4	ICV	1	Water	NA	NA	20.3	6.99	07/07/2020	08:15
5	CCV1	1	Water	NA	NA	20.1	2.02	07/07/2020	08:17
6	L3217-07	1	Water	NA	NA	24.5	1.37	07/07/2020	08:20
7	L3217-08	1	Water	NA	NA	25.1	13.34	07/07/2020	08:21
8	L3217-09	1	Water	NA	NA	24.7	12.54	07/07/2020	08:22
9	L3217-10	1	Water	NA	NA	24.7	8.50	07/07/2020	08:25
10	L3217-10DUP	1	Water	NA	NA	24.6	8.51	07/07/2020	08:26
11	CCV2	1	Water	NA	NA	20.1	12.02	07/07/2020	08:30

VB 109882

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PHWL3217 WorkList ID : 140192 Department : Wet-Chemistry Date : 07-07-2020 07:55:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/02/2020	Water	L3217-07-17	pH	Cool 4 deg C	ENVI77	P41	E3-41-ACID	07/02/2020	9040C
07/02/2020	Water	L3217-08-13	pH	Cool 4 deg C	ENVI77	P41	E3-41-BASE	07/02/2020	9040C
07/02/2020	Water	L3217-09-17	pH	Cool 4 deg C	ENVI77	P41	E3-41-NHL	07/02/2020	9040C
07/02/2020	Water	L3217-10-17	pH	Cool 4 deg C	ENVI77	P41	E3-41-CN-LIQ	07/02/2020	9040C

Date/Time 07/07/2020 8:00
 Received by: J.P. (Digestion Lab)
 Relinquished by:

Date/Time 07/07/2020 10:30
 Received by: J.P. (Digestion Lab)
 Relinquished by: J.P.