

LB 109893


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

41252



A= 0.0000e+000

B= 2.4302e-004

C= 2.9372e-002

Rho= 0.9998408

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	-81	0.000	-81				
0.2	0.200	0.207	0.007	729	0.0 %	729				
2.5	2.500	2.574	0.074	10471	0.0 %	10471				
5.0	5.000	4.869	-0.131	19913	0.0 %	19913				
7.5	7.500	7.487	-0.013	30688	0.0 %	30688				
10.0	10.000	10.054	0.054	41252	0.0 %	41252				

LB109883 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 07 Jul 2020 08:57:48							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type		
0.0 - 1	50	-81	-	0.0000	7470A	PPB	07 Jul 2020 09:14:47	S	Std		
0.2 - 1	50.2	729	-	0.2000	7470A	PPB	07 Jul 2020 09:16:55	S	Std		
2.5 - 1	52.5	10471	-	2.5000	7470A	PPB	07 Jul 2020 09:19:03	S	Std		
5.0 - 1	55	19913	-	5.0000	7470A	PPB	07 Jul 2020 09:21:10	S	Std		
7.5 - 1	57.5	30688	-	7.5000	7470A	PPB	07 Jul 2020 09:23:18	S	Std		
10.0 - 1	50	41252	-	10.0000	7470A	PPB	07 Jul 2020 09:35:03	S	Std		
ICV15 - 1	ICV15	16842	4.1222	-	7470A	PPB	07 Jul 2020 09:37:38	U	SMPL		
ICB15 - 1	ICB15	-137	-0.0039	-	7470A	PPB	07 Jul 2020 09:39:43	U	SMPL		
CCV46 - 1	CCV46	19456	4.7575	-	7470A	PPB	07 Jul 2020 09:41:49	U	SMPL		
CCB46 - 1	CCB46	-125	-0.0010	-	7470A	PPB	07 Jul 2020 09:43:55	U	SMPL		
CRA - 1	CRA	761	0.2143	-	7470A	PPB	07 Jul 2020 09:46:01	U	SMPL		
HighStd - 1	HighStd	41115	10.0209	-	7470A	PPB	07 Jul 2020 09:48:07	U	SMPL		
ChkStd - 1	ChkStd	27861	6.8000	-	7470A	PPB	07 Jul 2020 09:50:13	U	SMPL		
PB129968BL - 1	PBW	-144	-0.0056	-	7470A	PPB	07 Jul 2020 09:52:21	U	SMPL		
PB129968BS - 1	LCSW	14513	3.5563	-	7470A	PPB	07 Jul 2020 09:54:28	U	SMPL		
L3202-02 - 1	RO-COMP	-86	0.0085	-	7470A	PPB	07 Jul 2020 09:56:35	U	SMPL		
L3202-02DUP - 1	RO-COMPDUP	-23	0.0238	-	7470A	PPB	07 Jul 2020 09:58:43	U	SMPL		
L3202-02MS - 1	RO-COMPMS	15040	3.6843	-	7470A	PPB	07 Jul 2020 10:00:50	U	SMPL		
L3202-02MSD - 1	RO-COMPMSD	16156	3.9555	-	7470A	PPB	07 Jul 2020 10:02:58	U	SMPL		
L3204-02 - 1	WF-COMP	-89	0.0077	-	7470A	PPB	07 Jul 2020 10:05:05	U	SMPL		
CCV47 - 1	CCV47	19734	4.8250	-	7470A	PPB	07 Jul 2020 10:07:10	U	SMPL		
CCB47 - 1	CCB47	-110	0.0026	-	7470A	PPB	07 Jul 2020 10:09:16	U	SMPL		
L3205-02 - 1	12013	6	0.0308	-	7470A	PPB	07 Jul 2020 10:11:22	U	SMPL		
L3205-04 - 1	VNJ-214	-18	0.0250	-	7470A	PPB	07 Jul 2020 10:13:28	U	SMPL		
L3206-03 - 1	DRUM-COMP-2	-3	0.0286	-	7470A	PPB	07 Jul 2020 10:15:34	U	SMPL		
L3208-01 - 1	RT-4743	-8	0.0274	-	7470A	PPB	07 Jul 2020 10:17:41	U	SMPL		
PB129969BL - 1	PBW	-20	0.0245	-	7470A	PPB	07 Jul 2020 10:19:47	U	SMPL		
PB129969BS - 1	LCSW	16270	3.9832	-	7470A	PPB	07 Jul 2020 10:21:54	U	SMPL		
L3206-01 - 1	DRUM-COMP-1	-3	0.0286	-	7470A	PPB	07 Jul 2020 10:24:02	U	SMPL		
L3206-01DUP - 1	DRUM-COMP-1DUP	-11	0.0267	-	7470A	PPB	07 Jul 2020 10:26:10	U	SMPL		
L3206-01MS - 1	DRUM-COMP-1MS	14917	3.6544	-	7470A	PPB	07 Jul 2020 10:28:17	U	SMPL		
L3206-01MSD - 1	DRUM-COMP-1MSD	15608	3.8224	-	7470A	PPB	07 Jul 2020 10:30:24	U	SMPL		
CCV48 - 1	CCV48	19135	4.6795	-	7470A	PPB	07 Jul 2020 10:32:30	U	SMPL		
CCB48 - 1	CCB48	-113	0.0019	-	7470A	PPB	07 Jul 2020 10:34:37	U	SMPL		
PB129970BL - 1	PBW	-30	0.0221	-	7470A	PPB	07 Jul 2020 10:36:43	U	SMPL		
PB129970BS - 1	LCSW	14738	3.6109	-	7470A	PPB	07 Jul 2020 10:38:49	U	SMPL		
L3214-01 - 1	MW-14	-103	0.0043	-	7470A	PPB	07 Jul 2020 10:40:56	U	SMPL		
L3214-01DUP - 1	MW-14DUP	-40	0.0197	-	7470A	PPB	07 Jul 2020 10:43:02	U	SMPL		
L3214-01MS - 1	MW-14MS	14773	3.6194	-	7470A	PPB	07 Jul 2020 10:45:09	U	SMPL		
L3214-01MSD - 1	MW-14MSD	14838	3.6352	-	7470A	PPB	07 Jul 2020 10:47:16	U	SMPL		
L3214-02 - 1	SW-2	-99	0.0053	-	7470A	PPB	07 Jul 2020 10:49:24	U	SMPL		
L3214-03 - 1	MW-9R	152	0.0663	-	7470A	PPB	07 Jul 2020 10:51:31	U	SMPL		
L3214-04 - 1	MW-8R	23	0.0350	-	7470A	PPB	07 Jul 2020 10:53:39	U	SMPL		
L3214-05 - 1	SW-3	-23	0.0238	-	7470A	PPB	07 Jul 2020 10:55:45	U	SMPL		
CCV49 - 1	CCV49	19384	4.7400	-	7470A	PPB	07 Jul 2020 10:57:52	U	SMPL		
CCB49 - 1	CCB49	-107	0.0034	-	7470A	PPB	07 Jul 2020 10:59:58	U	SMPL		
L3214-06 - 1	SW-3-DUP	-42	0.0192	-	7470A	PPB	07 Jul 2020 11:02:05	U	SMPL		
PB129968TB - 1	PB129948TB	-4	0.0284	-	7470A	PPB	07 Jul 2020 11:04:11	U	SMPL		
PB129969TB - 1	PB129961TB	-24	0.0235	-	7470A	PPB	07 Jul 2020 11:06:18	U	SMPL		
L3202-02LX5 - 1		-36	0.0206	-	7470A	PPB	07 Jul 2020 11:08:24	U	SMPL		
L3202-02A - 1		15582	3.8160	-	7470A	PPB	07 Jul 2020 11:10:31	U	SMPL		
L3206-01LX5 - 1		-100	0.0051	-	7470A	PPB	07 Jul 2020 11:12:38	U	SMPL		
L3206-01A - 1		15929	3.9004	-	7470A	PPB	07 Jul 2020 11:14:46	U	SMPL		
L3214-01LX5 - 1		-105	0.0039	-	7470A	PPB	07 Jul 2020 11:16:54	U	SMPL		
L3214-01A - 1		16651	4.0758	-	7470A	PPB	07 Jul 2020 11:19:02	U	SMPL		
CCV50 - 1	CCV50	19161	4.6858	-	7470A	PPB	07 Jul 2020 11:21:08	U	SMPL		
CCB50 - 1	CCB50	-112	0.0022	-	7470A	PPB	07 Jul 2020 11:23:15	U	SMPL		