


LB109893

245.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	12	0.000	12				
0.05	0.050	0.063	0.013	253	0.0 %	253				
0.2	0.200	0.204	0.004	790	0.0 %	790				
2.5	2.500	2.484	-0.016	9455	0.0 %	9455				
5.0	5.000	4.936	-0.064	18772	0.0 %	18772				
7.5	7.500	7.620	0.120	28975	0.0 %	28975				
10.0	10.000	9.944	-0.056	37806	0.0 %	37806				

LB109893 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 07 Jul 2020 14:19:54

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	12	-	0.0000	245.1	PPB	07 Jul 2020 14:22:38	S	Std
0.05 - 1	50.5	253	-	0.0500	245.1	PPB	07 Jul 2020 14:24:55	S	Std
0.2 - 1	50.2	790	-	0.2000	245.1	PPB	07 Jul 2020 14:27:12	S	Std
2.5 - 1	52.5	9455	-	2.5000	245.1	PPB	07 Jul 2020 14:29:29	S	Std
5.0 - 1	55	18772	-	5.0000	245.1	PPB	07 Jul 2020 14:31:45	S	Std
7.5 - 1	57.5	28975	-	7.5000	245.1	PPB	07 Jul 2020 14:34:01	S	Std
10.0 - 1	510	37806	-	10.0000	245.1	PPB	07 Jul 2020 14:36:17	S	Std
ICV17 - 1	ICV17	14802	3.8910	-245.1		PPB	07 Jul 2020 14:39:04	U	SMPL
ICB17 - 1	ICB17	-67	-0.0213	-245.1		PPB	07 Jul 2020 14:41:18	U	SMPL
CCV54 - 1	CCV54	18977	4.9895	-245.1		PPB	07 Jul 2020 14:43:33	U	SMPL
CCB54 - 1	CCB54	-75	-0.0234	-245.1		PPB	07 Jul 2020 14:45:48	U	SMPL
CRA - 1	CRA	235	0.0581	-245.1		PPB	07 Jul 2020 14:48:04	U	SMPL
HighStd - 1	HighStd	37649	9.9024	-245.1		PPB	07 Jul 2020 14:50:19	U	SMPL
ChkStd - 1	ChkStd	26335	6.9255	-245.1		PPB	07 Jul 2020 14:52:34	U	SMPL
PB130013BL - 1	PBW	-139	-0.0403	-245.1		PPB	07 Jul 2020 14:54:52	U	SMPL
PB130013BS - 1	LCSW	13628	3.5821	-245.1		PPB	07 Jul 2020 14:57:10	U	SMPL
L2816-01 - 1	COMP	639	0.1644	-245.1		PPB	07 Jul 2020 14:59:26	U	SMPL
L2816-01DUP - 1	COMPDUP	706	0.1821	-245.1		PPB	07 Jul 2020 15:01:42	U	SMPL
L2816-01MS - 1	COMPMS	12461	3.2750	-245.1		PPB	07 Jul 2020 15:06:33	U	SMPL
L2816-01MSD - 1	COMPMSD	12443	3.2703	-245.1		PPB	07 Jul 2020 15:08:49	U	SMPL
L3194-01 - 1	TUNNEL-WALL-WASHWATER(EXHAUST)	1484	0.3868	-245.1		PPB	07 Jul 2020 15:11:05	U	SMPL
CCV55 - 1	CCV55	18170	4.7772	-245.1		PPB	07 Jul 2020 15:13:20	U	SMPL
CCB55 - 1	CCB55	-79	-0.0245	-245.1		PPB	07 Jul 2020 15:15:36	U	SMPL
L2816-01LX5 - 1		115	0.0266	-245.1		PPB	07 Jul 2020 15:17:52	U	SMPL
L2816-01A - 1		14430	3.7931	-245.1		PPB	07 Jul 2020 15:20:07	U	SMPL
CCV57 - 1	CCV57	18639	4.9006	-245.1		PPB	07 Jul 2020 15:22:22	U	SMPL
CCB57 - 1	CCB57	-70	-0.0221	-245.1		PPB	07 Jul 2020 15:24:39	U	SMPL

CCV56
CCB56
CCV57
CCB57
NB
7/13/20