



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109894

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	07/07/2020	08:10
2	CAL2	1	Water	NA	NA	20.1	7.00	07/07/2020	08:11
3	CAL3	1	Water	NA	NA	20.4	10.03	07/07/2020	08:12
4	ICV	1	Water	NA	NA	20.3	6.99	07/07/2020	08:15
5	CCV1	1	Water	NA	NA	20.1	2.02	07/07/2020	08:17
6	L3206-12	1	Solid	20.03	20	24.3	9.05	07/07/2020	08:35
7	L3206-12DUP	1	Solid	20.04	20	24.4	9.06	07/07/2020	08:36
8	L3217-01	1	Solid	20.02	20	24.8	1.56	07/07/2020	08:37
9	L3217-02	1	Solid	20.03	20	24.2	8.75	07/07/2020	08:40
10	L3217-03	1	Solid	20.03	20	24.0	9.87	07/07/2020	08:42
11	L3217-04	1	Solid	20.05	20	24.2	9.43	07/07/2020	08:43
12	L3217-05	1	Solid	20.03	20	24.6	7.15	07/07/2020	08:45
13	CCV2	1	Water	NA	NA	20.3	12.02	07/07/2020	08:50

✓ 10/08/20

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S L3217 WorkList ID : 140191 Department : Wet-Chemistry Date : 07-07-2020 07:54:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/02/2020	Solid	L3206-12	✓ pH	Cool 4 deg C	PSEG01	P33	31078	07/02/2020	9045D
07/02/2020	Solid	L3217-01	✓ pH	Cool 4 deg C	ENVI77	P41	E3-41-DRUM-95	07/02/2020	9045D
07/02/2020	Solid	L3217-02	✓ pH	Cool 4 deg C	ENVI77	P41	E3-41-NHS	07/02/2020	9045D
07/02/2020	Solid	L3217-03	✓ pH	Cool 4 deg C	ENVI77	P41	E3-41-COMP-1-FLOOR	07/02/2020	9045D
07/02/2020	Solid	L3217-04	✓ pH	Cool 4 deg C	ENVI77	P41	E3-41-VAT	07/02/2020	9045D
07/02/2020	Solid	L3217-05	✓ pH	Cool 4 deg C	ENVI77	P41	E3-41-BOILER-ROOM	07/02/2020	9045D

Date/Time 07/07/2020 8.00
 Received by: JO (WAT) (am)
 Relinquished by: KS (wetchem)

Date/Time 07/07/2020 10.25
 Received by: KS (wetchem)
 Relinquished by: JO (WAT) (am)