



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109896

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/07/2020	11:05
2	L3194-01		1	8	100.00	>212.0	>212.0	07/07/2020	11:35
3	L3194-01DUP		1	8	100.00	>212.0	>212.0	07/07/2020	11:55
4	L3206-01		1	9	100.00	>212.0	>212.0	07/07/2020	12:20
5	L3207-01		1	9	100.00	>212.0	>212.0	07/07/2020	12:50
6	L3217-09		1	8	100.00	>212.0	>212.0	07/07/2020	13:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 109895

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp7720 WorkList ID : 140215 Department : Wet-Chemistry Date : 07-07-2020 10:31:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/08/2020	Water	L3194-01	Flash Point	Cool 4 deg C	TULL02	P21	TUNNEL-WALL-WASHWATEF	07/01/2020	1010A
07/16/2020	Water	L3206-01	Flash Point	Cool 4 deg C	PSEG01	P33	DRUM-COMP-1	07/02/2020	1010A
07/07/2020	Water	L3207-01	Flash Point	Cool 4 deg C	PSEG01	P33	31063	07/02/2020	1010A
07/16/2020	Water	L3217-09	Flash Point	Cool 4 deg C	ENVI77	P41	E3-41-NHL	07/02/2020	1010A

Date/Time
07/07/2020 10:35

Received by:
KS (Wetchem)

Relinquished by:
JP (W11- (ntm))

Date/Time
07/07/2020 13:30

Received by:
JP (W11- (ntm))

Relinquished by:
KS (Wetchem)