



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109908

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/08/2020	11:30
2	CAL2	1	Water	NA	NA	20.2	7.00	07/08/2020	11:31
3	CAL3	1	Water	NA	NA	20.2	10.03	07/08/2020	11:32
4	ICV	1	Water	NA	NA	20.2	7.01	07/08/2020	11:35
5	CCV1	1	Water	NA	NA	20.2	2.01	07/08/2020	11:36
6	L3235-01	1	Water	NA	NA	15.3	6.58	07/08/2020	11:40
7	L3235-01DUP	1	Water	NA	NA	15.4	6.59	07/08/2020	11:41
8	L3235-03	1	Water	NA	NA	16.2	6.69	07/08/2020	11:45
9	CCV2	1	Water	NA	NA	20.1	12.02	07/08/2020	11:46

2B109908

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 13235 WorkList ID : 140247 Department : Wet-Chemistry Date : 07-08-2020 11:22:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/07/2020	Water	L3235-01- C	pH	Cool 4 deg C	APT101	P22	WGGM-20-WWW-TSF-022	07/07/2020	SM 4500-H B
07/07/2020	Water	L3235-03 -C	pH	Cool 4 deg C	APT101	P22	WGGM-20-WWW-TSF-023	07/07/2020	SM 4500-H B

Date/Time 07/08/2020 11:25
Received by: Jo (WST-Chem)
Relinquished by: [Signature]

Date/Time 07/08/2020 12:00
Received by: [Signature]
Relinquished by: Jo (WST-Chem)