


LB109912

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.054	0.054	368	0.000	368				
0.2	0.200	0.233	0.033	1293	0.0 %	1293				
2.5	2.500	2.484	-0.016	12880	0.0 %	12880				
5.0	5.000	4.890	-0.110	25264	0.0 %	25264				
7.5	7.500	7.436	-0.064	38370	0.0 %	38370				
10.0	10.000	10.104	0.104	52104	0.0 %	52104				

LB109912 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 08 Jul 2020 14:22:44					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	368	-	0.0000	7470A	PPB	08 Jul 2020 14:26:37	S	Std
0.2 - 1	S0.2	1293	-	0.2000	7470A	PPB	08 Jul 2020 14:36:51	S	Std
2.5 - 1	S2.5	12880	-	2.5000	7470A	PPB	08 Jul 2020 14:38:58	S	Std
5.0 - 1	S5	25264	-	5.0000	7470A	PPB	08 Jul 2020 14:41:06	S	Std
7.5 - 1	S7.5	38370	-	7.5000	7470A	PPB	08 Jul 2020 14:45:38	S	Std
10.0 - 1	S10	52104	-	10.0000	7470A	PPB	08 Jul 2020 14:50:10	S	Std
ICV19 - 1	ICV19	19399	3.7505	-	7470A	PPB	08 Jul 2020 14:56:22	U	SMPL
ICB19 - 1	ICB19	260	0.0326	-	7470A	PPB	08 Jul 2020 14:58:27	U	SMPL
CCV60 - 1	CCV60	25198	4.8769	-	7470A	PPB	08 Jul 2020 15:00:33	U	SMPL
CCB60 - 1	CCB60	262	0.0330	-	7470A	PPB	08 Jul 2020 15:02:39	U	SMPL
CRA - 1	CRA	1307	0.2360	-	7470A	PPB	08 Jul 2020 15:04:44	U	SMPL
HighStd - 1	HighStd	51933	10.0703	-	7470A	PPB	08 Jul 2020 15:06:50	U	SMPL
ChkStd - 1	ChkStd	35587	6.8951	-	7470A	PPB	08 Jul 2020 15:11:52	U	SMPL
PB130031BL - 1	PBW	172	0.0155	-	7470A	PPB	08 Jul 2020 15:13:58	U	SMPL
PB130031BS - 1	LCSW	17276	3.3381	-	7470A	PPB	08 Jul 2020 15:16:05	U	SMPL
L3217-15 - 1	E3-41-ACID	3898	0.7393	-	7470A	PPB	08 Jul 2020 15:18:12	U	SMPL
L3217-16 - 1	E3-41-BASE	199	0.0208	-	7470A	PPB	08 Jul 2020 15:20:19	U	SMPL
L3217-17 - 1	E3-41-NHL	5527	1.0558	-	7470A	PPB	08 Jul 2020 15:22:28	U	SMPL
L3217-18 - 1	E3-41-CN-LIQ	4038	0.7665	-	7470A	PPB	08 Jul 2020 15:24:36	U	SMPL
L3217-18DUP - 1	E3-41-CN-LIQDUP	3764	0.7133	-	7470A	PPB	08 Jul 2020 15:26:42	U	SMPL
CCV61 - 1	CCV61	25566	4.9484	-	7470A	PPB	08 Jul 2020 15:28:49	U	SMPL
CCB61 - 1	CCB61	279	0.0363	-	7470A	PPB	08 Jul 2020 15:30:55	U	SMPL
L3217-18MS - 1	E3-41-CN-LIQMS	20492	3.9628	-	7470A	PPB	08 Jul 2020 15:33:01	U	SMPL
L3217-18MSD - 1	E3-41-CN-LIQMSD	20403	3.9455	-	7470A	PPB	08 Jul 2020 15:37:39	U	SMPL
PB130032BL - 1	PBW	229	0.0266	-	7470A	PPB	08 Jul 2020 15:42:07	U	SMPL
PB130032BS - 1	LCSW	19005	3.6739	-	7470A	PPB	08 Jul 2020 15:44:13	U	SMPL
L3183-04 - 1	OK-02-070720	601	0.0989	-	7470A	PPB	08 Jul 2020 15:46:19	U	SMPL
L3187-04 - 1	SU-02-070720	598	0.0983	-	7470A	PPB	08 Jul 2020 15:48:26	U	SMPL
L3217-06 - 1	E3-41-DRUM-95	689	0.1160	-	7470A	PPB	08 Jul 2020 15:50:34	U	SMPL
L3217-11 - 1	E3-41-NHS	6614	1.2669	-	7470A	PPB	08 Jul 2020 15:52:41	U	SMPL
L3217-12 - 1	E3-41-COMP-1-FLOOR	2447	0.4575	-	7470A	PPB	08 Jul 2020 15:54:49	U	SMPL
L3217-13 - 1	E3-41-VAT	1517	0.2768	-	7470A	PPB	08 Jul 2020 15:56:55	U	SMPL
CCV62 - 1	CCV62	25366	4.9096	-	7470A	PPB	08 Jul 2020 15:59:02	U	SMPL
CCB62 - 1	CCB62	267	0.0340	-	7470A	PPB	08 Jul 2020 16:01:08	U	SMPL
L3217-14 - 1	E3-41-BOILER-ROOM	1740	0.3201	-	7470A	PPB	08 Jul 2020 16:03:14	U	SMPL
L3228-03 - 1	VNJ-217	1163	0.2080	-	7470A	PPB	08 Jul 2020 16:10:04	U	SMPL
L3228-03DUP - 1	VNJ-217DUP	1138	0.2032	-	7470A	PPB	08 Jul 2020 16:14:18	U	SMPL
L3228-03MS - 1	VNJ-217MS	20098	3.8862	-	7470A	PPB	08 Jul 2020 16:18:58	U	SMPL
L3228-03MSD - 1	VNJ-217MSD	20498	3.9639	-	7470A	PPB	08 Jul 2020 16:21:05	U	SMPL
PB130033BL - 1	PBW	5	-0.0169	-	7470A	PPB	08 Jul 2020 16:23:12	U	SMPL
PB130033BS - 1	LCSW	19044	3.6815	-	7470A	PPB	08 Jul 2020 16:25:19	U	SMPL
L3217-07 - 1	E3-41-ACID	7738	1.4853	-	7470A	PPB	08 Jul 2020 16:27:27	U	SMPL
L3217-08 - 1	E3-41-BASE	41205	7.9864	-	7470A	PPB	08 Jul 2020 16:31:17	U	SMPL
L3217-09 - 1	E3-41-NHL	43733	8.4774	-	7470A	PPB	08 Jul 2020 16:33:23	U	SMPL
CCV63 - 1	CCV63	26643	5.1576	-	7470A	PPB	08 Jul 2020 16:38:39	U	SMPL
CCB63 - 1	CCB63	-120	-0.0412	-	7470A	PPB	08 Jul 2020 16:40:45	U	SMPL
L3217-10 - 1	E3-41-CN-LIQ	29045	5.6242	-	7470A	PPB	08 Jul 2020 16:42:53	U	SMPL
L3217-10DUP - 1	E3-41-CN-LIQDUP	28256	5.4710	-	7470A	PPB	08 Jul 2020 16:44:59	U	SMPL
L3217-10MS - 1	E3-41-CN-LIQMS	34671	6.7171	-	7470A	PPB	08 Jul 2020 16:47:07	U	SMPL
L3217-10MSD - 1	E3-41-CN-LIQMSD	36993	7.1682	-	7470A	PPB	08 Jul 2020 16:49:13	U	SMPL
PB130032TB - 1	PB130014TB	30	-0.0120	-	7470A	PPB	08 Jul 2020 16:51:21	U	SMPL
PB130031TB - 1	PB130016TB	410	0.0618	-	7470A	PPB	08 Jul 2020 16:53:29	U	SMPL
CCV64 - 1	CCV64	25447	4.9253	-	7470A	PPB	08 Jul 2020 16:58:03	U	SMPL
CCB64 - 1	CCB64	37	-0.0107	-	7470A	PPB	08 Jul 2020 17:00:10	U	SMPL
L3217-18LX5 - 1		901	0.1571	-	7470A	PPB	08 Jul 2020 17:05:16	U	SMPL
L3217-18A - 1		23727	4.5912	-	7470A	PPB	08 Jul 2020 17:09:50	U	SMPL
L3228-03LX5 - 1		534	0.0859	-	7470A	PPB	08 Jul 2020 17:11:55	U	SMPL
L3228-03A - 1		21749	4.2070	-	7470A	PPB	08 Jul 2020 17:14:03	U	SMPL
L3217-10LX5 - 1		5341	1.0196	-	7470A	PPB	08 Jul 2020 17:18:35	U	SMPL
L3217-10A - 1		38036	7.3708	-	7470A	PPB	08 Jul 2020 17:20:41	U	SMPL
CCV65 - 1	CCV65	23930	4.6306	-	7470A	PPB	08 Jul 2020 17:22:47	U	SMPL
CCB65 - 1	CCB65	112	0.0039	-	7470A	PPB	08 Jul 2020 17:24:56	U	SMPL