



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109919

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/08/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	07/08/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	07/08/2020	17:12
4	ICV	1	Water	NA	NA	20.2	6.98	07/08/2020	17:15
5	CCV1	1	Water	NA	NA	20.1	2.02	07/08/2020	17:19
6	L3186-05	1	Solid	20.03	20	25.1	9.37	07/08/2020	17:20
7	L3186-05DUP	1	Solid	20.04	20	25.2	9.38	07/08/2020	17:21
8	L3203-03	1	Solid	20.02	20	25.3	6.30	07/08/2020	17:22
9	L3203-07	1	Solid	20.03	20	25.5	7.54	07/08/2020	17:25
10	L3203-11	1	Solid	20.04	20	25.5	8.53	07/08/2020	17:26
11	L3203-15	1	Solid	20.02	20	25.6	8.99	07/08/2020	17:29
12	L3203-19	1	Solid	20.03	20	25.3	7.90	07/08/2020	17:30
13	CCV2	1	Water	NA	NA	20.3	12.02	07/08/2020	17:32

WORKLIST(Hardcopy Internal Chain)

VB 109919

WorkList Name : CORROSIONITY L3203 WorkList ID : 140251 Department : Wet-Chemistry Date : 07-08-2020 11:50:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/08/2020	Solid	L3186-05	Corrosivity	Cool 4 deg C	PSEG05	P31	OR-02-070820	07/08/2020	9045D
07/07/2020	Solid	L3203-03	Corrosivity	Cool 4 deg C	PSEG01	P43	RO-24	07/07/2020	9045D
07/07/2020	Solid	L3203-07	Corrosivity	Cool 4 deg C	PSEG01	P43	RO-10	07/07/2020	9045D
07/07/2020	Solid	L3203-11	Corrosivity	Cool 4 deg C	PSEG01	P43	RO-20	07/07/2020	9045D
07/07/2020	Solid	L3203-15	Corrosivity	Cool 4 deg C	PSEG01	P43	RO-06	07/07/2020	9045D
07/07/2020	Solid	L3203-19	Corrosivity	Cool 4 deg C	PSEG01	P43	RO-16	07/07/2020	9045D

Date/Time 07/08/2020 16:40
 Received by: Jn (sub) (m)
 Relinquished by: Jn (sub) (m)

Date/Time 07/08/2020 18:10
 Received by: Jn (sub) (m)
 Relinquished by: Jn (sub) (m)