



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109924

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/09/2020	12:40
2	CAL2	1	Water	NA	NA	20.3	7.01	07/09/2020	12:41
3	CAL3	1	Water	NA	NA	20.3	10.03	07/09/2020	12:42
4	ICV	1	Water	NA	NA	20.3	7.00	07/09/2020	12:44
5	CCV1	1	Water	NA	NA	20.2	2.01	07/09/2020	12:45
6	L3250-03	1	Solid	20.03	20	24.0	6.33	07/09/2020	12:46
7	L3250-04	1	Solid	20.04	20	23.6	6.43	07/09/2020	12:48
8	L3250-05	1	Solid	20.02	20	23.7	6.92	07/09/2020	12:49
9	L3250-06	1	Solid	20.03	20	23.6	7.11	07/09/2020	12:50
10	L3250-07	1	Solid	20.02	20	23.6	6.98	07/09/2020	13:00
11	L3250-08	1	Solid	20.03	20	23.1	7.00	07/09/2020	13:01
12	L3250-09	1	Solid	20.03	20	23.7	7.28	07/09/2020	13:02
13	L3251-01	1	Solid	20.04	20	23.4	7.11	07/09/2020	13:03
14	L3251-04	1	Solid	20.02	20	23.6	8.16	07/09/2020	13:05
15	L3251-05	1	Solid	20.03	20	23.6	5.96	07/09/2020	13:06
16	CCV2	1	Water	NA	NA	20.1	12.02	07/09/2020	13:08
17	L3251-05DUP	1	Solid	20.04	20	23.7	5.97	07/09/2020	13:09
18	CCV3	1	Water	NA	NA	20.3	2.01	07/09/2020	13:10

18 109924

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3250 WorkList ID : 140290 Department : Wet-Chemistry Date : 07-09-2020 12:30:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/08/2020	Solid	L3250-03	pH	Cool 4 deg C	PSEG01	P51	31087	07/08/2020	9045D
07/08/2020	Solid	L3250-04	pH	Cool 4 deg C	PSEG01	P51	31056	07/08/2020	9045D
07/08/2020	Solid	L3250-05	pH	Cool 4 deg C	PSEG01	P51	31088	07/08/2020	9045D
07/08/2020	Solid	L3250-06	pH	Cool 4 deg C	PSEG01	P51	31090	07/08/2020	9045D
07/08/2020	Solid	L3250-07	pH	Cool 4 deg C	PSEG01	P51	31091	07/08/2020	9045D
07/08/2020	Solid	L3250-08	pH	Cool 4 deg C	PSEG01	P51	31092	07/08/2020	9045D
07/08/2020	Solid	L3250-09	pH	Cool 4 deg C	PSEG01	P51	31033	07/08/2020	9045D
07/08/2020	Solid	L3251-01	pH	Cool 4 deg C	PSEG01	P52	BUR-DRUM-COMP	07/08/2020	9045D
07/08/2020	Solid	L3251-04	pH	Cool 4 deg C	PSEG01	P52	31085	07/08/2020	9045D
07/08/2020	Solid	L3251-05	pH	Cool 4 deg C	PSEG01	P52	31131	07/08/2020	9045D

Date/Time 07/09/2020 12:35
 Received by: Jo [Signature] (mem)
 Relinquished by: Jo [Signature] (sm)

Date/Time 07/09/2020 16:00
 Received by: [Signature] (sm)
 Relinquished by: Jo [Signature] (mem)