



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB109925

Analyst By : JIGNESH

Supervisor Review By : apatel

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/09/2020	12:40
2	CAL2	1	Water	NA	NA	20.3	7.01	07/09/2020	12:41
3	CAL3	1	Water	NA	NA	20.3	10.03	07/09/2020	12:42
4	ICV	1	Water	NA	NA	20.3	7.00	07/09/2020	12:44
5	CCV1	1	Water	NA	NA	20.2	2.01	07/09/2020	12:45
6	L3250-12	1	Water	NA	NA	19.9	4.54	07/09/2020	13:15
7	L3250-14	1	Water	NA	NA	18.3	6.04	07/09/2020	13:16
8	L3250-15	1	Water	NA	NA	18.4	5.21	07/09/2020	13:17
9	L3250-16	1	Water	NA	NA	18.5	6.72	07/09/2020	13:19
10	L3250-17	1	Water	NA	NA	19.1	6.86	07/09/2020	13:20
11	L3250-18	1	Water	NA	NA	20.1	5.29	07/09/2020	13:22
12	L3250-19	1	Water	NA	NA	19.5	6.84	07/09/2020	13:25
13	L3250-20	1	Water	NA	NA	19.3	5.25	07/09/2020	13:26
14	L3250-21	1	Water	NA	NA	19.3	5.12	07/09/2020	13:27
15	L3250-22	1	Water	NA	NA	20.4	7.23	07/09/2020	13:30
16	CCV2	1	Water	NA	NA	20.3	12.02	07/09/2020	13:32
17	L3251-06	1	Water	NA	NA	21.4	8.87	07/09/2020	13:35
18	L3252-01	1	Water	NA	NA	22.3	8.97	07/09/2020	13:36
19	L3252-01DUP	1	Water	NA	NA	22.4	8.98	07/09/2020	13:37
20	CCV3	1	Water	NA	NA	20.1	2.02	07/09/2020	13:30

13109925

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH W L3250 WorkList ID : 140291 Department : Wet-Chemistry Date : 07-09-2020 12:31:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/08/2020	Water	L3250-12	pH	Cool 4 deg C	PSEG01	P51	31089	07/08/2020	9040C
07/08/2020	Water	L3250-14	pH	Cool 4 deg C	PSEG01	P51	31135	07/08/2020	9040C
07/08/2020	Water	L3250-15	pH	Cool 4 deg C	PSEG01	P51	31136	07/08/2020	9040C
07/08/2020	Water	L3250-16	pH	Cool 4 deg C	PSEG01	P51	31137	07/08/2020	9040C
07/08/2020	Water	L3250-17	pH	Cool 4 deg C	PSEG01	P51	31138	07/08/2020	9040C
07/08/2020	Water	L3250-18	pH	Cool 4 deg C	PSEG01	P51	31139	07/08/2020	9040C
07/08/2020	Water	L3250-19	pH	Cool 4 deg C	PSEG01	P51	31140	07/08/2020	9040C
07/08/2020	Water	L3250-20	pH	Cool 4 deg C	PSEG01	P51	31141	07/08/2020	9040C
07/08/2020	Water	L3250-21	pH	Cool 4 deg C	PSEG01	P51	31142	07/08/2020	9040C
07/08/2020	Water	L3250-22	pH	Cool 4 deg C	PSEG01	P51	31143	07/08/2020	9040C
07/08/2020	Water	L3251-06	pH	Cool 4 deg C	PSEG01	P52	31105	07/08/2020	9040C
07/08/2020	Water	L3252-01	pH	Cool 4 deg C	PSEG01	P52	TRE-1032	07/08/2020	9040C

Date/Time 07/09/2020 12:35
 Received by: JP (wet-chem)
 Relinquished by: CP (SM)

Date/Time 07/09/2020 14:00
 Received by: CP (SM)
 Relinquished by: JP (Wet-chem)