



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109932

Slope : 98.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	07/09/2020	14:20
2	CAL2	1	Water	NA	NA	20.4	7.01	07/09/2020	14:21
3	CAL3	1	Water	NA	NA	20.4	10.03	07/09/2020	14:22
4	ICV	1	Water	NA	NA	20.4	6.98	07/09/2020	14:25
5	CCV1	1	Water	NA	NA	20.3	2.01	07/09/2020	14:26
6	L3266-01	1	Water	NA	NA	26.7	5.03	07/09/2020	14:30
7	L3266-01DUP	1	Water	NA	NA	26.8	5.04	07/09/2020	14:31
8	L3266-02	1	Water	NA	NA	26.6	4.76	07/09/2020	14:32
9	CCV2	1	Water	NA	NA	20.3	12.02	07/09/2020	14:35

HB 109932

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3256 WorkList ID : 140297 Department : Wet-Chemistry Date : 07-09-2020 14:10:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/06/2020	Water	L3266-01	pH	Cool 4 deg C	STER16	P41	MAIN-CELL	07/06/2020	SM 4500-H B
07/06/2020	Water	L3266-02	pH	Cool 4 deg C	STER16	P41	EX-CELL	07/06/2020	SM 4500-H B

Date/Time 07/09/2020 14:15
Received by: Jo Twit (in m)
Relinquished by: AN Gwet chern

Date/Time 07/09/2020 15:20
Received by: AN Gwet chern
Relinquished by: Jo Twit (in m)