



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109936

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/09/2020	17:15
2	CAL2	1	Water	NA	NA	20.1	7.00	07/09/2020	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	07/09/2020	17:18
4	ICV	1	Water	NA	NA	20.2	7.00	07/09/2020	17:20
5	CCV1	1	Water	NA	NA	20.3	2.01	07/09/2020	17:21
6	L3184-06	1	Solid	20.03	20	25.6	9.14	07/09/2020	17:25
7	L3184-06DUP	1	Solid	20.04	20	25.5	9.15	07/09/2020	17:26
8	L3256-04	1	Solid	20.02	20	24.7	8.38	07/09/2020	17:29
9	L3259-02	1	Solid	20.03	20	25.3	5.55	07/09/2020	17:30
10	CCV2	1	Water	NA	NA	20.3	12.02	07/09/2020	17:32

18 109936

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3252 WorkList ID : 140300 Department : Wet-Chemistry Date : 07-09-2020 14:35:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/09/2020	Solid	L3184-06	Corrosivity	Cool 4 deg C	PSEG05	P53	EO-02-070920	07/09/2020	9045D
07/09/2020	Solid	L3256-04	Corrosivity	Cool 4 deg C	PSEG01	P52	24465	07/09/2020	9045D
07/09/2020	Solid	L3259-02	Corrosivity	Cool 4 deg C	PSEG01	P52	T-2-DRUM	07/09/2020	9045D

Date/Time 07092020 16:15
 Received by: J. CP (Wet)
 Relinquished by: J. CP (Sm)

(nm)

Date/Time 07092020 18:20
 Received by: CP (Sm)
 Relinquished by: J. CP (Wet)
 (nm)