

LB109944


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42824



A= 0.0000e+000

B= 2.3190e-004

C= -1.7586e-002

Rho= 0.9998386

Accept=Accepted

Slope
X-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.030	0.030	207	0.000	207					1
0.05	0.050	0.051	0.001	296	0.0 %	296					2
0.20	0.200	0.188	-0.012	887	0.0 %	887					-6
2.50	2.500	2.434	-0.066	10570	0.0 %	10570					-3
5.00	5.000	4.999	-0.001	21632	0.0 %	21632					0
7.50	7.500	7.635	0.135	32998	0.0 %	32998					2
10.0	10.000	9.913	-0.087	42824	0.0 %	42824					-1

LB109944

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.05	207	-		0 ISM02.4	PPB	7/10/2020 12:11	Std
	0.05 S0.05	296	-	0.05	ISM02.4	PPB	7/10/2020 12:16	Std
	0.2 S0.2	887	-	0.2	ISM02.4	PPB	7/10/2020 12:19	Std
	2.5 S2.5	10570	-	2.5	ISM02.4	PPB	7/10/2020 12:21	Std
	5 S5.0	21632	-	5	ISM02.4	PPB	7/10/2020 12:23	Std
	7.5 S7.5	32998	-	7.5	ISM02.4	PPB	7/10/2020 12:30	Std
	10 S10.0	42824	-	10	ISM02.4	PPB	7/10/2020 12:36	Std
ICV	ICV	17180	3.9665 -		ISM02.4	PPB	7/10/2020 12:38	SMPL
ICB	ICB	-8	-0.0194 -		ISM02.4	PPB	7/10/2020 12:41	SMPL
CCV075	CCV075	22226	5.1366 -		ISM02.4	PPB	7/10/2020 12:43	SMPL
CCB075	CCB075	67	-0.002 -		ISM02.4	PPB	7/10/2020 12:48	SMPL
PB130098BL	PBW098	50	-0.006 -		ISM02.4	PPB	7/10/2020 12:50	SMPL
L3241-01	MBFRZ8	-22	-0.0227 -		ISM02.4	PPB	7/10/2020 12:52	SMPL
L3241-02	MBFS14	5540	1.2671 -		ISM02.4	PPB	7/10/2020 12:54	SMPL
L3241-03	MBFS66	1956	0.436 -		ISM02.4	PPB	7/10/2020 12:57	SMPL
L3241-04	MBFS16	759	0.1584 -		ISM02.4	PPB	7/10/2020 12:59	SMPL
L3241-05	MBFS18	2986	0.6749 -		ISM02.4	PPB	7/10/2020 13:01	SMPL
L3241-06	MBFS18D	2604	0.5863 -		ISM02.4	PPB	7/10/2020 13:04	SMPL
L3241-07	MBFS18S	6975	1.5999 -		ISM02.4	PPB	7/10/2020 13:06	SMPL
PB130099BL	PBW099	-1	-0.0178 -		ISM02.4	PPB	7/10/2020 13:13	SMPL
L3242-01	MBFS15	76	0 -		ISM02.4	PPB	7/10/2020 13:16	SMPL
L3242-02	MBFS67	-37	-0.0262 -		ISM02.4	PPB	7/10/2020 13:18	SMPL
L3242-03	MBFS17	-118	-0.0449 -		ISM02.4	PPB	7/10/2020 13:20	SMPL
L3242-04	MBFS19	63	-0.003 -		ISM02.4	PPB	7/10/2020 13:22	SMPL
L3242-05	MBFS19D	46	-0.0069 -		ISM02.4	PPB	7/10/2020 13:25	SMPL
L3242-06	MBFS19S	3878	0.8817 -		ISM02.4	PPB	7/10/2020 13:27	SMPL
CCV076	CCV076	22172	5.1241 -		ISM02.4	PPB	7/10/2020 13:32	SMPL
CCB076	CCB076	-2	-0.018 -		ISM02.4	PPB	7/10/2020 13:36	SMPL