



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109945

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/10/2020	10:15
2	L3250-10		1	8	100.00	>212.0	>212.0	07/10/2020	10:40
3	L3250-10DUP		1	8	100.00	>212.0	>212.0	07/10/2020	11:10
4	L3250-12		1	9	100.00	>212.0	>212.0	07/10/2020	11:35
5	L3250-22		1	9	100.00	>212.0	>212.0	07/10/2020	12:05
6	L3251-06		1	9	100.00	>212.0	>212.0	07/10/2020	12:35
7	L3252-01		1	8	100.00	>212.0	>212.0	07/10/2020	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 109945

WorkList Name : fp71020 WorkList ID : 140333 Department : Wet-Chemistry Date : 07-10-2020 09:42:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/21/2020	Water	L3250-10 <i>E</i>	Flash Point	Cool 4 deg C	PSEG01	P51	DRUM-VAULT-WATER-COMF	07/08/2020	1010A
07/21/2020	Water	L3250-12 <i>B</i>	Flash Point	Cool 4 deg C	PSEG01	P51	31089	07/08/2020	1010A
07/21/2020	Water	L3250-22 <i>C</i>	Flash Point	Cool 4 deg C	PSEG01	P51	31143	07/08/2020	1010A
07/21/2020	Water	L3251-06 <i>B</i>	Flash Point	Cool 4 deg C	PSEG01	P52	31105	07/08/2020	1010A
07/21/2020	Water	L3252-01 <i>F</i>	Flash Point	Cool 4 deg C	PSEG01	P52	TRE-1032	07/08/2020	1010A

Date/Time 07/10/2020 09:45
 Received by: KS (wet-chem)
 Relinquished by: SP (VAULT-chem)

Date/Time 07/10/2020 13:15
 Received by: SP (wet-chem)
 Relinquished by: KS (wet-chem)