



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB109969

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP85400
Turbidity Calibration std, 0NTU	WP85399
Turbidity Calibration std, 20NTU	WP85395
Turbidity Calibration std, 40NTU	WP85394
Turbidity Calibration std, 1NTU	WP85398
Turbidity Calibration std, 5NTU	WP85397
Turbidity LOQ std, 1.0NTU	WP85401
Turbidity Calibration - CCV std, 10 NTU	WP85396
10 NTU Standard 500 ml	W2489

Intercept: 0.2184

Slope: 1.0606

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.266	07/13/2020	10:05
2	CAL2	1	1	1.330	07/13/2020	10:08
3	CAL3	5	1	5.620	07/13/2020	10:10
4	CAL4	10	1	10.600	07/13/2020	10:13
5	CAL5	20	1	21.400	07/13/2020	10:16
6	CAL6	40	1	42.700	07/13/2020	10:18



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB109969

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.300	9.506	07/13/2020	10:22
2	ICB		1	0.256	0.035	07/13/2020	10:25
3	CCV1	10	1	10.700	9.883	07/13/2020	10:28
4	CCB1		1	0.271	0.050	07/13/2020	10:31
5	LB109969BL		1	0.275	0.053	07/13/2020	10:34
6	L3216-10		1	0.801	0.549	07/13/2020	10:38
7	L3216-11		1	1.400	1.114	07/13/2020	10:41
8	CCV2	10	1	10.700	9.883	07/13/2020	10:44
9	CCB2		1	0.280	0.058	07/13/2020	10:47