



Analytical Summary Report

Analysis Method: SM4500 Cl G
Parameter: Residual Chlorine
Run Number: LB109971

ANALYST: Alexander
SUPERVISOR REVIEW BY: apatel

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP85387
Residual chlorine ICV-LCS, 0.4PPM	WP85383
Chlorine Calibration std, 0.1ppm	WP85381
Chlorine Calibration std, 0.2ppm	WP85382
Chlorine Calibration std, 0.8ppm	WP85384
Chlorine Calibration std, 0.0ppm	WP85380
Chlorine Calibration std, 1.6ppm	WP85385
Residual Chlorine LOQ 0.1ppm	WP85388
Residual Chlorine Calibration and CCV std, 0	WP85386
Total Chlorine Powder Pillows	W2249

Intercept: 0.0031 Slope: 1.0584 Regression: 0.999961

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	07/13/2020	11:05
2	CAL2	0.1	1	0.000	0.110	0.110	07/13/2020	11:08
3	CAL3	0.2	1	0.000	0.220	0.220	07/13/2020	11:11
4	CAL4	0.4	1	0.000	0.430	0.430	07/13/2020	11:14
5	CAL5	0.8	1	0.000	0.840	0.840	07/13/2020	11:16
6	CAL6	1.6	1	0.000	1.700	1.700	07/13/2020	11:18



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB109971

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4200	0.4200	0.3940	07/13/2020	11:20
2	ICB				1	0.0000	0.0000	0.0000	-0.0030	07/13/2020	11:22
3	CCV1			0.4	1	0.0000	0.4300	0.4300	0.4030	07/13/2020	11:24
4	CCB1				1	0.0000	0.0000	0.0000	-0.0030	07/13/2020	11:26
5	LB109971BL	50	50		1	0.0000	0.0000	0.0000	-0.0030	07/13/2020	11:28
6	LB109971BS	50	50	0.4	1	0.0000	0.4200	0.4200	0.3940	07/13/2020	11:30
7	L3216-07	50	50		1	0.0000	0.0500	0.0500	0.0440	07/13/2020	11:32
8	L3216-08	50	50		1	0.0000	0.1100	0.1100	0.1010	07/13/2020	11:36
9	CCV2			0.4	1	0.0000	0.4100	0.4100	0.3840	07/13/2020	11:38
10	CCB2				1	0.0000	0.0000	0.0000	-0.0030	07/13/2020	11:40