


LB109972

7470A

INSTRUMENT ID: CY1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.080	0.080	114	0.000	114				
0.2	0.200	0.242	0.042	656	0.0 %	656				
2.5	2.500	2.549	0.049	8397	0.0 %	8397				
5.0	5.000	4.741	-0.259	15753	0.0 %	15753				
7.5	7.500	7.432	-0.068	24780	0.0 %	24780				
10.0	10.000	10.157	0.157	33925	0.0 %	33925				

LB109972 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Jul 2020 11:05:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	114	-	0.0000	7470A	PPB	13 Jul 2020 11:11:27	S	Std
0.2 - 1	50.2	656	-	0.2000	7470A	PPB	13 Jul 2020 11:13:35	S	Std
2.5 - 1	52.5	8397	-	2.5000	7470A	PPB	13 Jul 2020 11:15:43	S	Std
5.0 - 1	55	15753	-	5.0000	7470A	PPB	13 Jul 2020 11:17:50	S	Std
7.5 - 1	57.5	24780	-	7.5000	7470A	PPB	13 Jul 2020 11:19:58	S	Std
10.0 - 1	51.1	33925	-	10.0000	7470A	PPB	13 Jul 2020 11:24:29	S	Std
ICV24 - 1	ICV24	13745	4.1426	-	7470A	PPB	13 Jul 2020 11:27:10	U	SMPL
ICB24 - 1	ICB24	-246	-0.0273	-	7470A	PPB	13 Jul 2020 11:29:16	U	SMPL
CCV83 - 1	CCV83	16153	4.8603	-	7470A	PPB	13 Jul 2020 11:31:21	U	SMPL
CCB83 - 1	CCB83	-251	-0.0288	-	7470A	PPB	13 Jul 2020 11:33:27	U	SMPL
CRA - 1	CRA	559	0.2126	-	7470A	PPB	13 Jul 2020 11:35:34	U	SMPL
HighStd - 1	HighStd	33781	10.1142	-	7470A	PPB	13 Jul 2020 11:37:39	U	SMPL
ChkStd - 1	ChkStd	23734	7.1198	-	7470A	PPB	13 Jul 2020 11:39:45	U	SMPL
PB130115BL - 1	PBW	-443	-0.0860	-	7470A	PPB	13 Jul 2020 11:41:53	U	SMPL
PB130115BS - 1	LCSW	12930	3.8997	-	7470A	PPB	13 Jul 2020 11:44:01	U	SMPL
L3184-06 - 1	EO-02-070920	-292	-0.0410	-	7470A	PPB	13 Jul 2020 11:46:08	U	SMPL
L3184-06DUP - 1	EO-02-070920DUP	-139	0.0046	-	7470A	PPB	13 Jul 2020 11:48:16	U	SMPL
L3184-06MS - 1	EO-02-070920MS	12618	3.8067	-	7470A	PPB	13 Jul 2020 11:50:23	U	SMPL
L3184-06MSD - 1	EO-02-070920MSD	13095	3.9489	-	7470A	PPB	13 Jul 2020 11:52:32	U	SMPL
L3251-02 - 1	BUR-DRUM-COMP	-321	-0.0496	-	7470A	PPB	13 Jul 2020 11:54:37	U	SMPL
CCV84 - 1	CCV84	16072	4.8362	-	7470A	PPB	13 Jul 2020 11:56:43	U	SMPL
CCB84 - 1	CCB84	-314	-0.0475	-	7470A	PPB	13 Jul 2020 11:58:50	U	SMPL
L3255-04 - 1	CHRT-33	-116	0.0115	-	7470A	PPB	13 Jul 2020 12:00:57	U	SMPL
L3256-03 - 1	351-337	-95	0.0177	-	7470A	PPB	13 Jul 2020 12:03:03	U	SMPL
L3256-04 - 1	24465	-42	0.0335	-	7470A	PPB	13 Jul 2020 12:05:09	U	SMPL
L3259-02 - 1	T-2-DRUM	-68	0.0258	-	7470A	PPB	13 Jul 2020 12:07:15	U	SMPL
PB130115TB - 1	PB130080TB	-81	0.0219	-	7470A	PPB	13 Jul 2020 12:09:22	U	SMPL
L3184-06LX5 - 1		-80	0.0222	-	7470A	PPB	13 Jul 2020 12:11:29	U	SMPL
L3184-06A - 1		12820	3.8669	-	7470A	PPB	13 Jul 2020 12:13:36	U	SMPL
CCV85 - 1	CCV85	15898	4.7843	-	7470A	PPB	13 Jul 2020 12:15:43	U	SMPL
CCB85 - 1	CCB85	-304	-0.0446	-	7470A	PPB	13 Jul 2020 12:17:52	U	SMPL