



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 7/14/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 16:40
In Date: 07/13/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 07/14/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB109979

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L1955-15	MDL-ME-SOIL-01	25	1.00	2.00	2.00	100.0	
L1955-16	MDL-ME-SOIL-02	26	1.00	2.00	2.00	100.0	
L1955-17	MDL-ME-SOIL-03	27	1.00	2.00	2.00	100.0	
L1955-18	MDL-ME-SOIL-04	28	1.00	2.00	2.00	100.0	
L1955-19	MDL-ME-SOIL-05	29	1.00	2.00	2.00	100.0	
L1955-20	MDL-ME-SOIL-06	30	1.00	2.00	2.00	100.0	
L1955-21	MDL-ME-SOIL-07	31	1.00	2.00	2.00	100.0	
L3179-01	HR-05-071320	45	1.14	9.72	9.14	93.2	
L3179-02	HR-05-071320	46	1.19	9.77	9.25	93.9	
L3179-03	HR-05-071320-1	47	1.16	9.97	9.5	94.7	
L3179-04	HR-05-071320-2	48	1.13	9.8	9.11	92.0	
L3276-01	WASTE	1	1.13	9.57	8.4	86.1	
L3276-02	VOC	2	1.13	9.73	8.68	87.8	
L3276-03	1	3	1.16	9.77	8.61	86.5	
L3276-04	2	4	1.18	9.98	8.72	85.7	
L3276-05	3	5	1.14	9.57	8.5	87.3	
L3276-06	4	6	1.11	9.83	8.59	85.8	
L3276-07	5	7	1.18	9.95	8.9	88.0	
L3282-01	ST-JOES-AVE-1	21	1.00	2.00	2.00	100.0	concrete SAMPLE,100% SOLIDS
L3282-02	ST-JOES-AVE-2	22	1.00	2.00	2.00	100.0	concrete SAMPLE,100% SOLIDS
L3282-03	INSIDE-GATE-OLDEN-AVE	23	1.00	2.00	2.00	100.0	concrete SAMPLE,100% SOLIDS
L3282-04	OUTSIDE-GATE-OLDEN-AVE	24	1.00	2.00	2.00	100.0	concrete SAMPLE,100% SOLIDS
L3283-01	PILE-COMP	32	1.00	2.00	2.00	100.0	concrete SAMPLE,100% SOLIDS
L3283-03	TP	33	1.15	9.95	8.7	85.8	
L3283-04	TP-EPH-2	34	1.00	2.00	2.00	100.0	concrete SAMPLE,100% SOLIDS
L3283-06	T-1	35	1.00	2.00	2.00	100.0	concrete SAMPLE,100% SOLIDS
L3283-08	T-2	36	1.00	2.00	2.00	100.0	concrete SAMPLE,100% SOLIDS
L3283-10	T-3	37	1.00	2.00	2.00	100.0	concrete SAMPLE,100% SOLIDS



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 7/14/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 16:40
In Date: 07/13/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 07/14/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB109979

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
L3283-12	FOUND-COMP	38	1.00	2.00	2.00	100.0	concrete SAMPLE,100% SOLIDS
L3283-14	SLAB-COMP	39	1.00	2.00	2.00	100.0	concrete SAMPLE,100% SOLIDS
L3284-01	T1-OIL	40	1.00	2.00	2.00	100.0	OIL SAMPLE,100% SOLIDS
L3285-01	S2-01-B	8	1.18	9.62	8.96	92.2	
L3285-02	S2-01-S	9	1.13	9.66	9.22	94.8	
L3285-03	S2-02-B	10	1.15	9.76	8.9	90.0	
L3285-04	S2-02-S	11	1.13	9.6	8.93	92.1	
L3285-05	S2-03-S	12	1.16	9.53	9.04	94.1	
L3285-06	S2-03-B	13	1.15	9.97	9.04	89.5	
L3285-07	S2-07-B	14	1.14	9.84	9.1	91.5	
L3285-08	S2-07-S	15	1.18	9.59	9.03	93.3	
L3285-09	S2-08-B	16	1.19	9.71	9.11	93.0	
L3285-10	S2-08-S	17	1.17	9.69	8.92	91.0	
L3285-11	S2-09-B	18	1.15	9.98	8.88	87.5	
L3285-12	S2-09-S	19	1.18	9.65	8.78	89.7	
L3285-13	DUP-3	20	1.18	9.94	8.79	86.9	
L3287-01	FK-GF-03-012-A	41	1.16	9.55	9.16	95.4	
L3287-02	FK-GF-03-012-A	42	1.18	9.79	9.33	94.7	
L3287-03	FK-GF-03-012-B	43	1.16	9.83	9.42	95.3	
L3287-04	FK-GF-03-012-B	44	1.18	9.72	9.27	94.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$