


LB109983

INSTRUMENT ID: CV4

245.1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	-68	0.000	-68				
0.05	0.050	0.058	0.008	157	0.0 %	157				
0.2	0.200	0.206	0.006	615	0.0 %	615				
2.5	2.500	2.498	-0.002	7724	0.0 %	7724				
5.0	5.000	5.027	0.027	15570	0.0 %	15570				
7.5	7.500	7.458	-0.042	23109	0.0 %	23109				
10.0	10.000	10.018	0.018	31049	0.0 %	31049				

LB109983 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 Jul 2020 13:27:40

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-68	-	0.0000	245.1	PPB	13 Jul 2020 13:34:50	S	Std
0.05 - 1	50.05	157	-	0.0500	245.1	PPB	13 Jul 2020 13:37:07	S	Std
0.2 - 1	50.2	615	-	0.2000	245.1	PPB	13 Jul 2020 13:39:25	S	Std
2.5 - 1	52.5	7724	-	2.5000	245.1	PPB	13 Jul 2020 13:41:41	S	Std
5.0 - 1	55	15570	-	5.0000	245.1	PPB	13 Jul 2020 13:43:58	S	Std
7.5 - 1	57.5	23109	-	7.5000	245.1	PPB	13 Jul 2020 13:46:13	S	Std
10.0 - 1	51.1	31049	-	10.0000	245.1	PPB	13 Jul 2020 13:51:41	S	Std
ICV25 - 1	ICV25	12171	3.9314	-245.1		PPB	13 Jul 2020 13:57:00	U	SMPL
ICB25 - 1	ICB25	-106	-0.0268	-245.1		PPB	13 Jul 2020 14:01:50	U	SMPL
CCV86 - 1	CCV86	16036	5.1776	-245.1		PPB	13 Jul 2020 14:04:05	U	SMPL
CCB86 - 1	CCB86	-96	-0.0236	-245.1		PPB	13 Jul 2020 14:08:54	U	SMPL
CRA - 1	CRA	105	0.0412	-245.1		PPB	13 Jul 2020 14:11:09	U	SMPL
HighStd - 1	HighStd	31742	10.2414	-245.1		PPB	13 Jul 2020 14:13:24	U	SMPL
ChkStd - 1	ChkStd	21728	7.0127	-245.1		PPB	13 Jul 2020 14:15:39	U	SMPL
PB130141BL - 1	PBW	-283	-0.0839	-245.1		PPB	13 Jul 2020 14:20:33	U	SMPL
PB130141BS - 1	LCSW	12617	4.0752	-245.1		PPB	13 Jul 2020 14:22:48	U	SMPL
L3233-02 - 1	SPRING-2020	1498	0.4903	-245.1		PPB	13 Jul 2020 14:29:58	U	SMPL
L3233-02DUP - 1	SPRING-2020DUP	1806	0.5896	-245.1		PPB	13 Jul 2020 14:37:47	U	SMPL
L3233-02MS - 1	SPRING-2020MS	12094	3.9066	-245.1		PPB	13 Jul 2020 14:53:47	U	SMPL
L3233-02MSD - 1	SPRING-2020MSD	12712	4.1059	-245.1		PPB	13 Jul 2020 14:56:04	U	SMPL
L3277-04 - 1	TI-633	-102	-0.0255	-245.1		PPB	13 Jul 2020 14:58:20	U	SMPL
CCV87 - 1	CCV87	14045	4.5356	-245.1		PPB	13 Jul 2020 15:00:35	U	SMPL
CCB87 - 1	CCB87	98	0.0390	-245.1		PPB	13 Jul 2020 15:02:50	U	SMPL
L3278-01 - 1	NCQ-632-EB	-128	-0.0339	-245.1		PPB	13 Jul 2020 15:05:05	U	SMPL
L3278-03 - 1	NCQ-632	86	0.0351	-245.1		PPB	13 Jul 2020 15:07:21	U	SMPL
L3278-04 - 1	NCQ-632-X	117	0.0451	-245.1		PPB	13 Jul 2020 15:09:36	U	SMPL
L3233-02LX5 - 1		199	0.0715	-245.1		PPB	13 Jul 2020 15:15:07	U	SMPL
L3233-02A - 1		13143	4.2448	-245.1		PPB	13 Jul 2020 15:17:23	U	SMPL
CCV88 - 1	CCV88	14600	4.7146	-245.1		PPB	13 Jul 2020 15:19:39	U	SMPL
CCB88 - 1	CCB88	74	0.0312	-245.1		PPB	13 Jul 2020 15:21:56	U	SMPL