



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109996

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/14/2020	10:40
2	CAL2	1	Water	NA	NA	20.3	6.98	07/14/2020	10:41
3	CAL3	1	Water	NA	NA	20.3	10.03	07/14/2020	10:42
4	ICV	1	Water	NA	NA	20.4	7.00	07/14/2020	10:44
5	CCV1	1	Water	NA	NA	20.3	2.02	07/14/2020	10:45
6	L3291-01	1	Water	NA	NA	15.5	5.49	07/14/2020	10:47
7	L3291-01DUP	1	Water	NA	NA	15.6	5.50	07/14/2020	10:48
8	L3291-02	1	Water	NA	NA	14.2	4.19	07/14/2020	10:50
9	L3291-03	1	Water	NA	NA	15.9	4.76	07/14/2020	10:51
10	CCV2	1	Water	NA	NA	20.3	12.02	07/14/2020	11:00

2B109996

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3291 WorkList ID : 140433 Department : Wet-Chemistry Date : 07-14-2020 10:34:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/13/2020	Water	L3291-01	pH	Cool 4 deg C	STER09	H12	CELL-A-POOL-WATER-(DE-IK)	07/13/2020	SM 4500-H B
07/13/2020	Water	L3291-02	pH	Cool 4 deg C	STER09	H12	CELL-B/C-POOL-WATER-(DE)	07/13/2020	SM 4500-H B
07/13/2020	Water	L3291-03	pH	Cool 4 deg C	STER09	H12	CELL-D-POOL-WATER-(DE-IK)	07/13/2020	SM 4500-H B

Date/Time 07-14-2020 10:35
Received by: Jo (wet chem)
Relinquished by: Jo (wet chem)

Date/Time 07-14-2020 11:40
Received by: AM (wet chem)
Relinquished by: Jo (wet chem)