



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB110000

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP85480
Turbidity Calibration std, 0NTU	WP85479
Turbidity Calibration std, 20NTU	WP85475
Turbidity Calibration std, 40NTU	WP85474
Turbidity Calibration std, 1NTU	WP85478
Turbidity Calibration std, 5NTU	WP85477
Turbidity Calibration - CCV std, 10 NTU	WP85476
10 NTU Standard 500 ml	W2489

Intercept: 0.1386

Slope: 1.057

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.212	07/14/2020	10:40
2	CAL2	1	1	1.270	07/14/2020	10:43
3	CAL3	5	1	5.480	07/14/2020	10:46
4	CAL4	10	1	10.500	07/14/2020	10:50
5	CAL5	20	1	21.200	07/14/2020	10:53
6	CAL6	40	1	42.500	07/14/2020	10:56



Analytical Summary Report

Analysis Method: 180.1
Parameter: Turbidity
Run Number: LB110000

ANALYST: Alexander
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.900	10.181	07/14/2020	11:02
2	ICB		1	0.265	0.120	07/14/2020	11:06
3	CCV1	10	1	10.400	9.708	07/14/2020	11:09
4	CCB1		1	0.284	0.138	07/14/2020	11:12
5	LB110000BL		1	0.272	0.126	07/14/2020	11:15
6	L3216-09		1	0.793	0.619	07/14/2020	11:19
7	CCV2	10	1	10.200	9.519	07/14/2020	11:22
8	CCB2		1	0.290	0.143	07/14/2020	11:25