



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB110003

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP85490
Residual chlorine ICV-LCS, 0.4PPM	WP85489
Chlorine Calibration std, 0.1ppm	WP85484
Chlorine Calibration std, 0.2ppm	WP85485
Chlorine Calibration std, 0.8ppm	WP85487
Chlorine Calibration std, 0.0ppm	WP85483
Chlorine Calibration std, 1.6ppm	WP85488
Residual Chlorine Calibration and CCV std, 0	WP85486
Total Chlorine Powder Pillows	W2249

Intercept: 0.0043

Slope: 1.0498

Regression: 0.999970

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	07/14/2020	12:10
2	CAL2	0.1	1	0.000	0.110	0.110	07/14/2020	12:12
3	CAL3	0.2	1	0.000	0.210	0.210	07/14/2020	12:15
4	CAL4	0.4	1	0.000	0.430	0.430	07/14/2020	12:18
5	CAL5	0.8	1	0.000	0.850	0.850	07/14/2020	12:21
6	CAL6	1.6	1	0.000	1.680	1.680	07/14/2020	12:23



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB110003

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4200	0.4200	0.3960	07/14/2020	12:26
2	ICB				1	0.0000	0.0000	0.0000	-0.0040	07/14/2020	12:29
3	CCV1			0.4	1	0.0000	0.4200	0.4200	0.3960	07/14/2020	12:32
4	CCB1				1	0.0000	0.0000	0.0000	-0.0040	07/14/2020	12:34
5	LB110003BL	50	50		1	0.0000	0.0000	0.0000	-0.0040	07/14/2020	12:36
6	LB110003BS	50	50	0.4	1	0.0000	0.4300	0.4300	0.4060	07/14/2020	12:39
7	L3216-09	50	50		1	0.0000	0.0600	0.0600	0.0530	07/14/2020	12:41
8	CCV2			0.4	1	0.0000	0.4200	0.4200	0.3960	07/14/2020	12:44
9	CCB2				1	0.0000	0.0000	0.0000	-0.0040	07/14/2020	12:46