


LB110009

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.8247e-004

C= 3.2898e-002

Rho= 0.9996646

Accept=Accepted

Slope  
Y=intercept

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.00   | 0.000  | 0.010  | 0.010  | -80   | 0.000      | -80   |       |       |       |       | 1  |
| 0.05   | 0.050  |        |        |       |            | 296   |       |       |       |       | 16 |
| 0.20   | 0.200  | 0.232  | 0.032  | 704   | 0.0 %      | 704   |       |       |       |       | 3  |
| 2.50   | 2.500  | 2.583  | 0.083  | 9027  | 0.0 %      | 9027  |       |       |       |       | 4  |
| 5.00   | 5.000  | 4.793  | -0.207 | 16853 | 0.0 %      | 16853 |       |       |       |       | 0  |
| 7.50   | 7.500  | 7.521  | 0.021  | 26511 | 0.0 %      | 26511 |       |       |       |       | 1  |
| 10.0   | 10.000 | 10.060 | 0.060  | 35500 | 0.0 %      | 35500 |       |       |       |       |    |

LB110009  
INSTRUMENT ID : CV1

| Sample ID  | Extended ID | $\mu$ Abs. | Conc.     | Std Conc | Method      | Units | Date            | Type |
|------------|-------------|------------|-----------|----------|-------------|-------|-----------------|------|
|            | 0 S0.0      | -80        | -         |          | 0 ISM02.4   | PPB   | 7/14/2020 15:05 | Std  |
|            | 0.2 S0.2    | 704        | -         |          | 0.2 ISM02.4 | PPB   | 7/14/2020 15:08 | Std  |
|            | 2.5 S2.5    | 9027       | -         |          | 2.5 ISM02.4 | PPB   | 7/14/2020 15:10 | Std  |
|            | 5 S5.0      | 16853      | -         |          | 5 ISM02.4   | PPB   | 7/14/2020 15:12 | Std  |
|            | 7.5 S7.5    | 26511      | -         |          | 7.5 ISM02.4 | PPB   | 7/14/2020 15:14 | Std  |
|            | 10 S10.0    | 35500      | -         |          | 10 ISM02.4  | PPB   | 7/14/2020 15:22 | Std  |
| ICV        | ICV         | 13936      | 3.9694 -  |          | ISM02.4     | PPB   | 7/14/2020 15:25 | SMPL |
| ICB        | ICB         | -309       | -0.0544 - |          | ISM02.4     | PPB   | 7/14/2020 15:30 | SMPL |
| CCV098     | CCV098      | 16862      | 4.7959 -  |          | ISM02.4     | PPB   | 7/14/2020 15:33 | SMPL |
| CCB098     | CCB098      | -235       | -0.0335 - |          | ISM02.4     | PPB   | 7/14/2020 15:37 | SMPL |
| PB130205BL | PBW205      | -81        | 0.01 -    |          | ISM02.4     | PPB   | 7/14/2020 15:40 | SMPL |
| L3293-01   | MBG098      | -51        | 0.0185 -  |          | ISM02.4     | PPB   | 7/14/2020 15:42 | SMPL |
| L3293-02   | MBG098D     | -67        | 0.014 -   |          | ISM02.4     | PPB   | 7/14/2020 15:44 | SMPL |
| L3293-03   | MBG098S     | 3612       | 1.0532 -  |          | ISM02.4     | PPB   | 7/14/2020 15:46 | SMPL |
| CCV099     | CCV099      | 16732      | 4.7591 -  |          | ISM02.4     | PPB   | 7/14/2020 15:49 | SMPL |
| CCB099     | CCB099      | -412       | -0.0835 - |          | ISM02.4     | PPB   | 7/14/2020 15:51 | SMPL |