



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110023

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/14/2020	16:25
2	CAL2	1	Water	NA	NA	20.3	7.01	07/14/2020	16:26
3	CAL3	1	Water	NA	NA	20.3	10.03	07/14/2020	16:27
4	ICV	1	Water	NA	NA	20.3	6.98	07/14/2020	16:30
5	CCV1	1	Water	NA	NA	20.3	2.01	07/14/2020	16:32
6	L3289-02	1	Solid	20.03	20	23.2	7.39	07/14/2020	16:33
7	L3289-02DUP	1	Solid	20.04	20	23.3	7.40	07/14/2020	16:35
8	L3289-07	1	Solid	20.02	20	23.1	7.93	07/14/2020	16:39
9	CCV2	1	Water	NA	NA	20.3	12.02	07/14/2020	16:40

2B110023

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 13289 WorkList ID : 140444 Department : Wet-Chemistry Date : 07-14-2020 13:31:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/14/2020	Solid	L3289-02	pH	Cool 4 deg C	PSEG01	H12	IRON-1	07/14/2020	9045D
07/14/2020	Solid	L3289-07	pH	Cool 4 deg C	PSEG01	H12	IRON-2	07/14/2020	9045D

Date/Time 07-16-2020 16:20
 Received by: JA (Wet-Chem)
 Relinquished by: CP (sm)

Date/Time 07-16-2020 18:20
 Received by: CP (sm)
 Relinquished by: JA (Wet-Chem)