



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110028

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.00	07/15/2020	11:48
2	CAL2	1	Water	NA	NA	20.1	7.00	07/15/2020	11:49
3	CAL3	1	Water	NA	NA	20.2	10.03	07/15/2020	11:50
4	ICV	1	Water	NA	NA	20.3	7.02	07/15/2020	12:00
5	CCV1	1	Water	NA	NA	20.3	2.00	07/15/2020	12:01
6	L3312-01	1	Water	NA	NA	24.2	5.32	07/15/2020	12:05
7	L3312-01DUP	1	Water	NA	NA	24.3	5.33	07/15/2020	12:06
8	CCV2	1	Water	NA	NA	20.3	12.02	07/15/2020	12:10

YB110028

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3312 WorkList ID : 140480 Department : Wet-Chemistry Date : 07-15-2020 11:41:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/14/2020	Water	L3312-01	pH	Cool 4 deg C	STER01	H31	POOL-WATER	07/14/2020	SM 4500-H B

Date/Time 07-15-2020 11:45
 Received by: Jo (wetchem)
 Relinquished by: AM (wetchem)

Date/Time 07-15-2020 12:35
 Received by: AM (wetchem)
 Relinquished by: Jo (wetchem)