

LB110038


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34170



A= 0.0000e+000

B= 2.9214e-004

C= 1.3512e-003

Rho= 0.9998412

Accept=Accepted

slope
+ 1.3512e-003

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.009	0.009	26	0.000	26					-10
0.05	0.050	0.045	-0.005	151	0.0 %	151					-4
0.20	0.200	0.193	-0.007	655	0.0 %	655					3
2.50	2.500	2.570	0.070	8791	0.0 %	8791					-3
5.00	5.000	4.865	-0.135	16648	0.0 %	16648					0
7.50	7.500	7.584	0.084	25957	0.0 %	25957					
10.0	10.000	9.984	-0.016	34170	0.0 %	34170					

LB110038
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	26	-		0 ISM02.4	PPB	7/15/2020 14:28	Std
	0.05 S0.05	151	-	0.05	ISM02.4	PPB	7/15/2020 14:33	Std
	0.2 S0.2	655	-	0.2	ISM02.4	PPB	7/15/2020 14:35	Std
	2.5 S2.5	8791	-	2.5	ISM02.4	PPB	7/15/2020 14:38	Std
	5 S5.0	16648	-	5	ISM02.4	PPB	7/15/2020 14:40	Std
	7.5 S7.5	25957	-	7.5	ISM02.4	PPB	7/15/2020 14:45	Std
	10 S10.0	34170	-	10	ISM02.4	PPB	7/15/2020 15:01	Std
ICV	ICV	13549	3.9596 -		ISM02.4	PPB	7/15/2020 15:04	SMPL
ICB	ICB	-71	-0.0194 -		ISM02.4	PPB	7/15/2020 15:14	SMPL
CCV005	CCV005	16798	4.9088 -		ISM02.4	PPB	7/15/2020 15:16	SMPL
CCB005	CCB005	-54	-0.0144 -		ISM02.4	PPB	7/15/2020 15:21	SMPL
PB130222BL	PBW222	7	0.0034 -		ISM02.4	PPB	7/15/2020 15:23	SMPL
L3238-01	MEW0R7	2003	0.5865 -		ISM02.4	PPB	7/15/2020 15:25	SMPL
L3238-02	MEW0R8	-178	-0.0507 -		ISM02.4	PPB	7/15/2020 15:32	SMPL
L3238-03	MEW0S0	276	0.082 -		ISM02.4	PPB	7/15/2020 15:35	SMPL
L3238-04	MEW0S0D	233	0.0694 -		ISM02.4	PPB	7/15/2020 15:40	SMPL
L3238-05	MEW0S0S	3890	1.1378 -		ISM02.4	PPB	7/15/2020 15:42	SMPL
CCV006	CCV006	15921	4.6526 -		ISM02.4	PPB	7/15/2020 15:44	SMPL
CCB006	CCB006	6	0.0031 -		ISM02.4	PPB	7/15/2020 15:49	SMPL