



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110050

Slope : 97.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/16/2020	11:05
2	CAL2	1	Water	NA	NA	20.3	7.01	07/16/2020	11:06
3	CAL3	1	Water	NA	NA	20.3	10.03	07/16/2020	11:07
4	ICV	1	Water	NA	NA	20.3	6.99	07/16/2020	11:10
5	CCV1	1	Water	NA	NA	20.3	2.01	07/16/2020	11:11
6	L3326-01	1	Water	NA	NA	20.1	5.26	07/16/2020	11:15
7	L3326-01DUP	1	Water	NA	NA	20.2	5.27	07/16/2020	11:16
8	L3327-01	1	Water	NA	NA	22.4	5.31	07/16/2020	11:20
9	CCV2	1	Water	NA	NA	20.3	12.03	07/16/2020	11:22

1B110050

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3126 WorkList ID : 140524 Department : Wet-Chemistry Date : 07-16-2020 10:51:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/10/2020	Water	L3326-01-A	pH	Cool 4 deg C	STER10	H51	CELL-A -POOL-WATER(DE-IC)	07/10/2020	SM 4500-H B
07/15/2020	Water	L3327-01-A	pH	Cool 4 deg C	STER12	H31	CELL-POOL-WATER-3RD-Q	07/15/2020	SM 4500-H B

Date/Time 07-16-2020 11:00
Received by: So Cuelb-Chen
Relinquished by: AM wet chem

Date/Time 07-16-2020 12:10
Received by: AM wet chem
Relinquished by: So Cuelb
Chen