


LB110054

7470A

INSTRUMENT ID: CV2



A= 0.0000e+000

B= 3.2219e-004

C= 2.3167e-002

Rho= 0.9998487

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.057	0.057	104	0.000	104				
0.2	0.200	0.245	0.045	690	0.0 %	690				
2.5	2.500	2.430	-0.070	7470	0.0 %	7470				
5.0	5.000	4.892	-0.108	15113	0.0 %	15113				
7.5	7.500	7.530	0.030	23300	0.0 %	23300				
10.0	10.000	10.045	0.045	31106	0.0 %	31106				

LB110054 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 16 Jul 2020 15:08:47				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	104	-	0.00007470A	PPB	16 Jul 2020 15:12:08	S	Std
0.2 - 1	50.2	690	-	0.20007470A	PPB	16 Jul 2020 15:14:16	S	Std
2.5 - 1	52.5	7470	-	2.50007470A	PPB	16 Jul 2020 15:16:24	S	Std
5.0 - 1	55	15113	-	5.00007470A	PPB	16 Jul 2020 15:18:32	S	Std
7.5 - 1	57.5	23300	-	7.50007470A	PPB	16 Jul 2020 15:23:34	S	Std
10.0 - 1	510	31106	-	10.00007470A	PPB	16 Jul 2020 15:25:40	S	Std
ICV31 - 1	ICV31	11405	3.6978	-7470A	PPB	16 Jul 2020 15:30:32	U	SMPL
ICB31 - 1	ICB31	-227	-0.0500	-7470A	PPB	16 Jul 2020 15:32:37	U	SMPL
CCV11 - 1	CCV11	14746	4.7742	-7470A	PPB	16 Jul 2020 15:34:43	U	SMPL
CCB11 - 1	CCB11	-255	-0.0590	-7470A	PPB	16 Jul 2020 15:36:49	U	SMPL
CRA - 1	CRA	660	0.2358	-7470A	PPB	16 Jul 2020 15:38:56	U	SMPL
HighStd - 1	HighStd	31378	10.1329	-7470A	PPB	16 Jul 2020 15:41:02	U	SMPL
ChkStd - 1	ChkStd	20999	6.7889	-7470A	PPB	16 Jul 2020 15:43:08	U	SMPL
PB130257BL - 1	PBW	0	0.0232	-7470A	PPB	16 Jul 2020 15:48:19	U	SMPL
PB130257BS - 1	LCSW	11159	3.6185	-7470A	PPB	16 Jul 2020 15:50:26	U	SMPL
L3183-17 - 1	OK-03-071520	-219	-0.0474	-7470A	PPB	16 Jul 2020 15:52:34	U	SMPL
L3183-17DUP - 1	OK-03-071520DUP	-53	0.0061	-7470A	PPB	16 Jul 2020 15:54:42	U	SMPL
L3183-17MS - 1	OK-03-071520MS	12974	4.2033	-7470A	PPB	16 Jul 2020 15:56:50	U	SMPL
L3183-17MSD - 1	OK-03-071520MSD	12425	4.0264	-7470A	PPB	16 Jul 2020 15:58:58	U	SMPL
L3184-12 - 1	EO-01-071520	-258	-0.0600	-7470A	PPB	16 Jul 2020 16:01:04	U	SMPL
CCV12 - 1	CCV12	15329	4.9620	-7470A	PPB	16 Jul 2020 16:03:11	U	SMPL
CCB12 - 1	CCB12	-243	-0.0551	-7470A	PPB	16 Jul 2020 16:05:17	U	SMPL
L3315-04 - 1	TP-16-S	-64	0.0025	-7470A	PPB	16 Jul 2020 16:07:24	U	SMPL
L3315-08 - 1	TP-16-D	-18	0.0174	-7470A	PPB	16 Jul 2020 16:09:30	U	SMPL
PB130257TB - 1	PB130231TB	19	0.0293	-7470A	PPB	16 Jul 2020 16:11:37	U	SMPL
L3183-17LX5 - 1		3	0.0241	-7470A	PPB	16 Jul 2020 16:18:36	U	SMPL
L3183-17A - 1		12396	4.0170	-7470A	PPB	16 Jul 2020 16:20:42	U	SMPL
CCV13 - 1	CCV13	14993	4.8538	-7470A	PPB	16 Jul 2020 16:22:49	U	SMPL
CCB13 - 1	CCB13	-238	-0.0535	-7470A	PPB	16 Jul 2020 16:24:57	U	SMPL