



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110062

Slope : 97.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/16/2020	16:10
2	CAL2	1	Water	NA	NA	20.2	7.00	07/16/2020	16:11
3	CAL3	1	Water	NA	NA	20.2	10.03	07/17/2020	16:12
4	ICV	1	Water	NA	NA	20.2	7.00	07/16/2020	16:15
5	CCV1	1	Water	NA	NA	20.3	2.02	07/16/2020	16:16
6	L3323-03	1	Water	NA	NA	22.6	6.66	07/16/2020	16:35
7	L3323-03DUP	1	Water	NA	NA	22.7	6.67	07/16/2020	16:36
8	L3324-05	1	Water	NA	NA	23.7	6.26	07/16/2020	16:37
9	CCV2	1	Water	NA	NA	20.3	12.02	07/16/2020	16:40

2B118062

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3333 WorkList ID : 140535 Department : Wet-Chemistry Date : 07-16-2020 15:55:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/16/2020	Water	L3323-03	pH	Cool 4 deg C	PSEG01	H21	223-224	07/16/2020	9040C
07/16/2020	Water	L3324-05	pH	Cool 4 deg C	PSEG01	H21	TOTE-50-DIAZ	07/16/2020	9040C

Date/Time 07/16/2020 16:00
Received by: Jo (wft)
Relinquished by: CP (sm)

Date/Time 07-16-2020 17:35
Received by: CP (sm)
Relinquished by: Jo (wft) (htm)