



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110063

Slope : 97.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/16/2020	16:10
2	CAL2	1	Water	NA	NA	20.2	7.00	07/16/2020	16:11
3	CAL3	1	Water	NA	NA	20.2	10.03	07/16/2020	16:12
4	ICV	1	Water	NA	NA	20.2	7.00	07/16/2020	16:15
5	CCV1	1	Water	NA	NA	20.3	2.02	07/16/2020	16:16
6	L3188-16	1	Solid	20.03	20	22.4	7.63	07/16/2020	16:45
7	L3335-04	1	Solid	20.02	20	23.8	8.05	07/16/2020	16:46
8	L3335-08	1	Solid	20.03	20	23.9	5.75	07/16/2020	16:49
9	L3335-12	1	Solid	20.04	20	23.6	7.45	07/16/2020	16:50
10	L3335-12DUP	1	Solid	20.03	20	23.5	7.46	07/16/2020	17:00
11	CCV2	1	Water	NA	NA	20.3	12.02	07/16/2020	17:02

2B110063

WORKLIST(Hardcopy Internal Chain)

Date: 07-16-2020 13:36:43

Department: Wet-Chemistry

WorkList ID: 140529

WorkList Name: CORROSIONITY L3190

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/16/2020	Solid	L3188-16	Corrosivity	Cool 4 deg C	PSEG05	H21	TR-04-071620	07/16/2020	9045D
07/16/2020	Solid	L3335-04	Corrosivity	Cool 4 deg C	PSEG01	H21	TP-12-S	07/16/2020	9045D
07/16/2020	Solid	L3335-08	Corrosivity	Cool 4 deg C	PSEG01	H21	TP-11-S	07/16/2020	9045D
07/16/2020	Solid	L3335-12	Corrosivity	Cool 4 deg C	PSEG01	H21	TP-11-D	07/16/2020	9045D

Date/Time 07-16-2020 16:00
Received by: Jp (yrf) (h-m)
Relinquished by: cp (sm)

Date/Time 07-16-2020 17:35
Received by: cp (sm)
Relinquished by: Jp (yrf) (h-m)