



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110076

Slope : 97.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	07/17/2020	10:35
2	CAL2	1	Water	NA	NA	20.1	7.00	07/17/2020	10:36
3	CAL3	1	Water	NA	NA	20.1	10.02	07/17/2020	10:37
4	ICV	1	Water	NA	NA	20.1	7.01	07/17/2020	10:38
5	CCV1	1	Water	NA	NA	20.2	2.01	07/17/2020	10:40
6	L3338-01	1	Water	NA	NA	23.7	4.70	07/17/2020	10:42
7	L3338-01DUP	1	Water	NA	NA	23.6	4.71	07/17/2020	10:43
8	CCV2	1	Water	NA	NA	20.3	12.02	07/17/2020	10:45

213110076

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3338

WorkList ID : 140555

Department : Wet-Chemistry

Date : 07-17-2020 10:25:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/15/2020	Water	L3338-01-1A	pH	Cool 4 deg C	STER14	H41	MINI-CELL	07/15/2020	SM 4500-H B

Date/Time 07-17-2020 10:27
Received by: Jo (wet-chem)
Relinquished by: APLwetchem

Date/Time 07-17-2020 11:30
Received by: APL(wetchem)
Relinquished by: Jo (wet-chem) 12:00