


LB110094

7471B

INSTRUMENT ID: CV4



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.030	-0.030	-25	0.000	-25				
0.2	0.200	0.217	0.017	1173	0.0 %	1173				
2.5	2.500	2.488	-0.012	12174	0.0 %	12174				
5.0	5.000	5.059	0.059	24628	0.0 %	24628				
7.5	7.500	7.474	-0.026	36325	0.0 %	36325				
10.0	10.000	9.992	-0.008	48518	0.0 %	48518				

LB110094 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Jul 2020 08:56:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-25	-	0.0000	7471B	PPB	20 Jul 2020 09:00:41	S	Std
0.2 - 1	50.2	1173	-	0.2000	7471B	PPB	20 Jul 2020 09:10:55	S	Std
2.5 - 1	52.5	12174	-	2.5000	7471B	PPB	20 Jul 2020 09:13:12	S	Std
5.0 - 1	55	24628	-	5.0000	7471B	PPB	20 Jul 2020 09:15:29	S	Std
7.5 - 1	57.5	36325	-	7.5000	7471B	PPB	20 Jul 2020 09:17:45	S	Std
10.0 - 1	57.8	48518	-	10.0000	7471B	PPB	20 Jul 2020 09:20:01	S	Std
ICV32 - 1	ICV32	19075	3.9129	-	7471B	PPB	20 Jul 2020 09:22:46	U	SMPL
ICB32 - 1	ICB32	-247	-0.0762	-	7471B	PPB	20 Jul 2020 09:25:02	U	SMPL
CCV14 - 1	CCV14	25204	5.1783	-	7471B	PPB	20 Jul 2020 09:27:18	U	SMPL
CCB14 - 1	CCB14	-270	-0.0810	-	7471B	PPB	20 Jul 2020 09:29:34	U	SMPL
CRA - 1	CRA	1155	0.2132	-	7471B	PPB	20 Jul 2020 09:31:50	U	SMP
HighStd - 1	HighStd	48989	10.0889	-	7471B	PPB	20 Jul 2020 09:34:05	U	SMPL
ChkStd - 1	ChkStd	35290	7.2606	-	7471B	PPB	20 Jul 2020 09:36:20	U	SMPL
PB130247BL - 1	PBS	-71	-0.0399	-	7471B	PPB	20 Jul 2020 09:38:38	U	SMPL
PB130247BS - 1	LCSS	20597	4.2271	-	7471B	PPB	20 Jul 2020 09:40:57	U	SMPL
L3308-01 - 1	1	2021	0.3920	-	7471B	PPB	20 Jul 2020 09:43:13	U	SMPL
L3308-01DUP - 1	1DUP	2407	0.4717	-	7471B	PPB	20 Jul 2020 09:45:30	U	SMPL
L3308-01MS - 1	1MS	21064	4.3236	-	7471B	PPB	20 Jul 2020 09:47:48	U	SMPL
L3308-01MSD - 1	1MSD	19673	4.0364	-	7471B	PPB	20 Jul 2020 09:50:05	U	SMPL
L3308-02 - 1	2	1280	0.2390	-	7471B	PPB	20 Jul 2020 09:52:22	U	SMPL
CCV15 - 1	CCV15	24979	5.1318	-	7471B	PPB	20 Jul 2020 09:54:40	U	SMPL
CCB15 - 1	CCB15	-223	-0.0713	-	7471B	PPB	20 Jul 2020 09:56:55	U	SMPL
L3308-03 - 1	3	1633	0.3119	-	7471B	PPB	20 Jul 2020 09:59:12	U	SMPL
L3308-04 - 1	4	1933	0.3738	-	7471B	PPB	20 Jul 2020 10:01:27	U	SMPL
L3308-05 - 1	5	1612	0.3076	-	7471B	PPB	20 Jul 2020 10:03:42	U	SMPL
L3308-06 - 1	6	2556	0.5025	-	7471B	PPB	20 Jul 2020 10:05:58	U	SMPL
L3308-07 - 1	7	2167	0.4221	-	7471B	PPB	20 Jul 2020 10:08:14	U	SMPL
L3308-08 - 1	8	1427	0.2694	-	7471B	PPB	20 Jul 2020 10:10:31	U	SMPL
L3308-09 - 1	9	1745	0.3350	-	7471B	PPB	20 Jul 2020 10:12:47	U	SMPL
L3308-10 - 1	10	1780	0.3422	-	7471B	PPB	20 Jul 2020 10:15:04	U	SMPL
L3308-11 - 1	11	1776	0.3414	-	7471B	PPB	20 Jul 2020 10:17:21	U	SMPL
L3308-12 - 1	12	1950	0.3773	-	7471B	PPB	20 Jul 2020 10:19:37	U	SMPL
CCV16 - 1	CCV16	25107	5.1583	-	7471B	PPB	20 Jul 2020 10:21:52	U	SMPL
CCB16 - 1	CCB16	-162	-0.0587	-	7471B	PPB	20 Jul 2020 10:24:08	U	SMPL
L3308-13 - 1	13	1927	0.3726	-	7471B	PPB	20 Jul 2020 10:26:25	U	SMPL
L3308-14 - 1	14	2213	0.4316	-	7471B	PPB	20 Jul 2020 10:28:40	U	SMPL
PB130248BL - 1	PBS	-99	-0.0457	-	7471B	PPB	20 Jul 2020 10:30:56	U	SMPL
PB130248BS - 1	LCSS	19886	4.0804	-	7471B	PPB	20 Jul 2020 10:33:12	U	SMPL
L3183-11 - 1	OK-03-071520	4620	0.9286	-	7471B	PPB	20 Jul 2020 10:35:27	U	SMPL
L3184-07 - 1	EO-01-071520	4507	0.9053	-	7471B	PPB	20 Jul 2020 10:37:45	U	SMPL
L3308-15 - 1	15	2385	0.4672	-	7471B	PPB	20 Jul 2020 10:40:03	U	SMPL
L3308-16 - 1	16	3309	0.6579	-	7471B	PPB	20 Jul 2020 10:42:21	U	SMPL
L3308-17 - 1	17	3125	0.6199	-	7471B	PPB	20 Jul 2020 10:44:38	U	SMPL
L3308-18 - 1	18	2986	0.5912	-	7471B	PPB	20 Jul 2020 10:46:53	U	SMPL
CCV17 - 1	CCV17	24675	5.0691	-	7471B	PPB	20 Jul 2020 10:49:09	U	SMPL
CCB17 - 1	CCB17	-113	-0.0486	-	7471B	PPB	20 Jul 2020 10:51:25	U	SMPL
L3308-19 - 1	19	4863	0.9788	-	7471B	PPB	20 Jul 2020 10:53:41	U	SMPL
L3308-20 - 1	20	3683	0.7351	-	7471B	PPB	20 Jul 2020 10:55:56	U	SMPL
L3308-21 - 1	21	4220	0.8460	-	7471B	PPB	20 Jul 2020 10:58:12	U	SMPL
L3308-22 - 1	22	2598	0.5111	-	7471B	PPB	20 Jul 2020 11:00:28	U	SMPL
L3308-23 - 1	23	2381	0.4663	-	7471B	PPB	20 Jul 2020 11:02:45	U	SMPL
L3308-24 - 1	24	4324	0.8675	-	7471B	PPB	20 Jul 2020 11:05:01	U	SMPL
L3310-01 - 1	S-1	4137	0.8289	-	7471B	PPB	20 Jul 2020 11:07:18	U	SMPL
L3315-03 - 1	TP-16-S	2638	0.5194	-	7471B	PPB	20 Jul 2020 11:09:35	U	SMPL
L3315-07 - 1	TP-16-D	966	0.1742	-	7471B	PPB	20 Jul 2020 11:11:52	U	SMPL
L3315-07DUP - 1	TP-16-DDUP	1028	0.1870	-	7471B	PPB	20 Jul 2020 11:14:08	U	SMPL
CCV18 - 1	CCV18	24859	5.1071	-	7471B	PPB	20 Jul 2020 11:16:23	U	SMPL
CCB18 - 1	CCB18	-118	-0.0496	-	7471B	PPB	20 Jul 2020 11:18:39	U	SMPL
L3315-07MS - 1	TP-16-DMS	22752	4.6721	-	7471B	PPB	20 Jul 2020 11:20:54	U	SMPL
L3315-07MSD - 1	TP-16-DMSD	22668	4.6547	-	7471B	PPB	20 Jul 2020 11:23:10	U	SMPL
L3308-01LX5 - 1		250	0.0264	-	7471B	PPB	20 Jul 2020 11:25:27	U	SMPL
L3308-01A - 1		23150	4.7542	-	7471B	PPB	20 Jul 2020 11:27:44	U	SMPL
L3315-07LX5 - 1		113	-0.0019	-	7471B	PPB	20 Jul 2020 11:30:01	U	SMPL
L3315-07A - 1		22626	4.6461	-	7471B	PPB	20 Jul 2020 11:32:19	U	SMPL
CCV19 - 1	CCV19	24723	5.0790	-	7471B	PPB	20 Jul 2020 11:34:36	U	SMPL
CCB19 - 1	CCB19	-180	-0.0624	-	7471B	PPB	20 Jul 2020 11:36:54	U	SMPL