



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 7/21/2020

OVENTEMP IN Celsius(°C): 108
Time IN: 16:30
In Date: 07/20/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:45
Out Date: 07/21/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB110105

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L3184-13	EO-03-072020	15	1.18	9.71	9.25	94.6	
L3184-14	EO-03-072020	16	1.18	9.9	9.33	93.5	
L3184-15	EO-03-072020-1	17	1.12	9.92	9.42	94.3	
L3184-16	EO-03-072020-2	18	1.19	9.82	9.24	93.3	
L3333-14	S4-01-S-02	1	1.15	9.63	8.51	86.8	
L3345-02	DB-1.2	2	1.19	9.87	8.49	84.1	
L3350-01	HH-TRANSITE	3	1.00	2.00	2.00	100.0	concrete SAMPLE,100% SOLIDS
L3351-01	SAMPLE-1A	4	1.18	9.78	9.02	91.2	
L3351-02	SAMPLE-1A	5	1.17	9.79	8.95	90.3	
L3351-03	SAMPLE-1B	6	1.19	9.63	9.23	95.3	
L3351-04	SAMPLE-1B	7	1.12	9.74	9.3	94.9	
L3352-01	HS-01	8	1.15	9.94	9.07	90.1	
L3354-01	BACK-YARD	9	1.15	9.93	9.32	93.1	
L3355-01	11977	10	1.16	7.93	7.44	92.8	
L3355-02	11977	11	1.14	9.97	9.25	91.8	
L3355-03	11935	12	1.19	9.96	5.65	50.9	
L3355-04	11935	13	1.13	9.92	8.33	81.9	
L3356-01	S5-04-B-01	14	1.13	9.81	8.8	88.4	
L3360-01	SP-1A	19	1.15	9.67	8.78	89.6	
L3360-02	SP-1B	20	1.11	9.67	8.45	85.7	
L3360-03	SP-2	21	1.14	9.95	8.62	84.9	
L3360-04	SP-3	22	1.13	9.79	8.82	88.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$