



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110111

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/20/2020	16:35
2	CAL2	1	Water	NA	NA	20.3	7.01	07/20/2020	16:36
3	CAL3	1	Water	NA	NA	20.3	10.03	07/20/2020	16:37
4	ICV	1	Water	NA	NA	20.4	6.98	07/20/2020	16:40
5	CCV1	1	Water	NA	NA	20.3	2.01	07/20/2020	16:41
6	L3184-17	1	Solid	20.03	20	26.9	9.10	07/20/2020	16:42
7	L3352-01	1	Solid	20.04	20	26.9	7.41	07/20/2020	16:44
8	L3352-01DUP	1	Solid	20.03	20	27.0	7.41	07/20/2020	16:45
9	CCV2	1	Water	NA	NA	20.3	12.02	07/20/2020	16:47

18110111

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3352

WorkList ID : 140615

Department : Wet-Chemistry

Date : 07-20-2020 14:35:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/20/2020	Solid	L3184-17	Corrosivity	Cool 4 deg C	PSEG05		EO-03-072020	07/20/2020	9045D
07/20/2020	Solid	L3352-01	Corrosivity	Cool 4 deg C	PSEG01	H23	HS-01	07/20/2020	9045D

Date/Time 07/20/2020 15:40
 Received by: Jo (wet-chem)
 Relinquished by: Al (sm)

Date/Time 07/20/2020 17:00
 Received by: CF (sm)
 Relinquished by: Jo (wet-chem)