

LB110119

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by :

Instrument ID : Konelab

7/21/2020 13:50

Test: ^{gpc} Total CN *Myk*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.641	0.0	0.073	
ICB1	0.756	0.0	0.002	
CCV1	241.682	0.0	0.183	
CCB1	0.742	0.0	0.002	
PB130346BL	0.529	0.0	0.002	
L3184-17	0.544	0.0	0.002	
L3184-17DUP	0.540	0.0	0.002	
L3352-01	0.722	0.0	0.002	
CCV2	240.587	0.0	0.182	
CCB2	0.951	0.0	0.002	

N	10
Mean	58.269
SD	100.8242
CV%	173.03

Aquakem v. 7.2AQ1

Results from time period:

Tue Jul 21 13:12:25 2020

Tue Jul 21 13:12:34 2020

DK

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time
0.OPPBCN	A	REACTIVE CN	P	-0.195	µg/l	7/21/2020 11:37:40
5.OPPBCN	A	REACTIVE CN	P	3.8017	µg/l	7/21/2020 11:37:41
10PPBCN	A	REACTIVE CN	P	10.9412	µg/l	7/21/2020 11:37:42
50PPBCN	A	REACTIVE CN	P	50.8581	µg/l	7/21/2020 11:37:43
100PPBCN	A	REACTIVE CN	P	100.4718	µg/l	7/21/2020 11:37:44
250PPBCN	A	REACTIVE CN	P	248.6185	µg/l	7/21/2020 11:37:45
500PPBCN	A	REACTIVE CN	P	500.5037	µg/l	7/21/2020 11:37:46
ICV1	S	REACTIVE CN	P	95.6415	µg/l	7/21/2020 13:12:25
ICB1	S	REACTIVE CN	P	0.7564	µg/l	7/21/2020 13:12:26
CCV1	S	REACTIVE CN	P	241.6817	µg/l	7/21/2020 13:12:27
CCB1	S	REACTIVE CN	P	0.7422	µg/l	7/21/2020 13:12:28
PB130346BL	S	REACTIVE CN	P	0.5287	µg/l	7/21/2020 13:12:29
L3184-17	S	REACTIVE CN	P	0.5438	µg/l	7/21/2020 13:12:30
L3184-17DUP	S	REACTIVE CN	P	0.5399	µg/l	7/21/2020 13:12:31
L3352-01	S	REACTIVE CN	P	0.7221	µg/l	7/21/2020 13:12:32
CCV2	S	REACTIVE CN	P	240.5874	µg/l	7/21/2020 13:12:33
CCB2	S	REACTIVE CN	P	0.951	µg/l	7/21/2020 13:12:34

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : 12

Instrument ID : Konelab

7/21/2020 13:42


Test Total CN

Accepted 7/21/2020 11:38

Factor 1336
Bias 0.002

Coeff. of det. 0.999973

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	-0.1950	0.0000	
2	5.0PPBCN	0.004	3.8017	5.0000	
3	10PPBCN	0.010	10.9412	10.0000	
4	50PPBCN	0.040	50.8581	50.0000	
5	100PPBCN	0.077	100.4718	100.0000	
6	250PPBCN	0.188	248.6185	250.0000	
7	500PPBCN	0.376	500.5037	500.0000	