


LB110121

ISM02.4

INSTRUMENT ID: CV2



A= 0.0000e+000

B= 2.7413e-004

C= -2.8367e-002

Rho= 0.9999499

Accept = Accepted

*slope
y=1m/accept*

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.023	-0.023	19	0.000	19					2
0.05	0.050	0.051	0.001	289	0.0 %	289					15
0.20	0.200	0.230	0.030	941	0.0 %	941					-2
2.50	2.500	2.446	-0.054	9026	0.0 %	9026					1
5.00	5.000	5.055	0.055	18542	0.0 %	18542					0
7.50	7.500	7.530	0.030	27573	0.0 %	27573					0
10.0	10.000	9.962	-0.038	36444	0.0 %	36444					0

LB110121
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.05	19	-		0 ISM02.4	PPB	7/21/2020 15:10	Std
	0.05 S0.05	289	-	0.05	ISM02.4	PPB	7/21/2020 15:12	Std
	0.2 S0.2	941	-	0.2	ISM02.4	PPB	7/21/2020 15:15	Std
	2.5 S2.5	9026	-	2.5	ISM02.4	PPB	7/21/2020 15:17	Std
	5 S5.0	18542	-	5	ISM02.4	PPB	7/21/2020 15:19	Std
	7.5 S7.5	27573	-	7.5	ISM02.4	PPB	7/21/2020 15:21	Std
	10 S10.0	36444	-	10	ISM02.4	PPB	7/21/2020 15:24	Std
ICV	ICV	13778	3.7486	-	ISM02.4	PPB	7/21/2020 15:27	SMPL
ICB	ICB	131	0.0075	-	ISM02.4	PPB	7/21/2020 15:32	SMPL
CCV031	CCV031	18435	5.0252	-	ISM02.4	PPB	7/21/2020 15:34	SMPL
CCB031	CCB031	131	0.0075	-	ISM02.4	PPB	7/21/2020 15:39	SMPL
PB130347BL	PBW347	79	-0.0067	-	ISM02.4	PPB	7/21/2020 15:44	SMPL
L3279-01	MBFS36	766	0.1816	-	ISM02.4	PPB	7/21/2020 15:46	SMPL
L3279-02	MBFS38	15448	4.2064	-	ISM02.4	PPB	7/21/2020 15:48	SMPL
L3279-03	MBFS32	4373	1.1704	-	ISM02.4	PPB	7/21/2020 15:51	SMPL
L3279-04	MBFS34	3799	1.0131	-	ISM02.4	PPB	7/21/2020 15:53	SMPL
L3279-05	MBFS42	5098	1.3691	-	ISM02.4	PPB	7/21/2020 15:55	SMPL
L3279-06	MBFS40	435	0.0909	-	ISM02.4	PPB	7/21/2020 16:05	SMPL
L3279-07	MBFS40D	369	0.0728	-	ISM02.4	PPB	7/21/2020 16:08	SMPL
L3279-08	MBFS40S	4390	1.1751	-	ISM02.4	PPB	7/21/2020 16:15	SMPL
L3279-09	MBFS20	5912	1.5923	-	ISM02.4	PPB	7/21/2020 16:17	SMPL
L3279-10	MBFS22	2569	0.6759	-	ISM02.4	PPB	7/21/2020 16:20	SMPL
L3279-11	MBFS24	3508	0.9333	-	ISM02.4	PPB	7/21/2020 16:22	SMPL
CCV032	CCV032	17206	4.6883	-	ISM02.4	PPB	7/21/2020 16:24	SMPL
CCB032	CCB032	114	0.0029	-	ISM02.4	PPB	7/21/2020 16:29	SMPL
PB130348BL	PBW348	121	0.0048	-	ISM02.4	PPB	7/21/2020 16:31	SMPL
L3280-01	MBFS37	1	-0.0281	-	ISM02.4	PPB	7/21/2020 16:34	SMPL
L3280-02	MBFS39	9	-0.0259	-	ISM02.4	PPB	7/21/2020 16:36	SMPL
L3280-03	MBFS33	1262	0.3176	-	ISM02.4	PPB	7/21/2020 16:38	SMPL
L3280-04	MBFS35	309	0.0563	-	ISM02.4	PPB	7/21/2020 16:41	SMPL
L3280-05	MBFS43	48	-0.0152	-	ISM02.4	PPB	7/21/2020 16:43	SMPL
L3280-06	MBFS21	197	0.0256	-	ISM02.4	PPB	7/21/2020 16:45	SMPL
L3280-07	MBFS41	-32	-0.0371	-	ISM02.4	PPB	7/21/2020 16:47	SMPL
L3280-08	MBFS41D	-3	-0.0292	-	ISM02.4	PPB	7/21/2020 16:50	SMPL
L3280-09	MBFS41S	4303	1.1512	-	ISM02.4	PPB	7/21/2020 16:52	SMPL
L3280-10	MBFS23	538	0.1191	-	ISM02.4	PPB	7/21/2020 16:54	SMPL
L3280-11	MBFS25	114	0.0029	-	ISM02.4	PPB	7/21/2020 16:56	SMPL
CCV033	CCV033	17207	4.6886	-	ISM02.4	PPB	7/21/2020 16:59	SMPL
CCB033	CCB033	125	0.0059	-	ISM02.4	PPB	7/21/2020 17:04	SMPL
PB130349BL	PBW349	116	0.0034	-	ISM02.4	PPB	7/21/2020 17:06	SMPL
L3339-01	MBFS26	1716	0.442	-	ISM02.4	PPB	7/21/2020 17:13	SMPL
L3339-02	MBFS26D	1925	0.4993	-	ISM02.4	PPB	7/21/2020 17:21	SMPL
L3339-03	MBFS26S	5282	1.4196	-	ISM02.4	PPB	7/21/2020 17:26	SMPL
L3339-04	MBFS28	765	0.1813	-	ISM02.4	PPB	7/21/2020 17:28	SMPL
L3339-05	MBFS30	8885	2.4073	-	ISM02.4	PPB	7/21/2020 17:30	SMPL
PB130350BL	PBW350	154	0.0138	-	ISM02.4	PPB	7/21/2020 17:35	SMPL

LB110121

INSTRUMENT ID : CV1

L3340-01	MBFS27	16	-0.024 -	ISM02.4	PPB	7/21/2020 17:37 SMPL
L3340-02	MBFS27D	17	-0.0237 -	ISM02.4	PPB	7/21/2020 17:39 SMPL
L3340-03	MBFS27S	3842	1.0248 -	ISM02.4	PPB	7/21/2020 17:42 SMPL
L3340-04	MBFS29	-23	-0.0347 -	ISM02.4	PPB	7/21/2020 17:44 SMPL
L3340-05	MBFS31	249	0.0399 -	ISM02.4	PPB	7/21/2020 17:46 SMPL
CCV034	CCV034	17495	4.7675 -	ISM02.4	PPB	7/21/2020 17:49 SMPL
CCB034	CCB034	54	-0.0136 -	ISM02.4	PPB	7/21/2020 17:51 SMPL