



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB110123

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP85707
Turbidity Calibration std, 20NTU	WP85703
Turbidity Calibration std, 40NTU	WP85702
Turbidity Calibration std, 1NTU	WP85706
Turbidity Calibration std, 5NTU	WP85705
Turbidity Calibration - CCV std, 10 NTU	WP85704

Intercept: 0.4684

Slope: 0.8945

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.212	07/21/2020	12:20
2	CAL2	1	1	1.290	07/21/2020	12:23
3	CAL3	5	1	5.180	07/21/2020	12:26
4	CAL4	10	1	9.610	07/21/2020	12:29
5	CAL5	20	1	18.300	07/21/2020	12:32
6	CAL6	40	1	36.200	07/21/2020	12:35



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB110123

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.860	10.499	07/21/2020	12:40
2	ICB		1	0.229	-0.268	07/21/2020	12:42
3	CCV1	10	1	9.710	10.332	07/21/2020	12:45
4	CCB1		1	0.218	-0.280	07/21/2020	12:47
5	LB110123BL		1	0.237	-0.259	07/21/2020	12:49
6	L3319-02		1	5.300	5.401	07/21/2020	12:52
7	CCV2	10	1	9.800	10.432	07/21/2020	12:54
8	CCB2		1	0.227	-0.270	07/21/2020	12:56

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-71620 **WorkList ID :** 140507 **Department :** Wet-Chemistry **Date :** 07-16-2020 09:01:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/15/2020	Water	L3319-01	Turbidity	Cool 4 deg C	CHEM02	QA Office	PT-TURB-WS	07/13/2020	180.1
07/15/2020	Water	L3319-02	Turbidity	Cool 4 deg C	CHEM02	QA Office	PT-TURB-WS	07/13/2020	SM2130 B

Date/Time	<u>7/12/2020</u>	<u>12:05</u>
Received by:	<u>AM(wetchem)</u>	<u>AF (wetchem)</u>
Relinquished by:	<u>AF (wetchem)</u>	<u>AM(wetchem)</u>