



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110128

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	24.7	4.01	07/21/2020	16:35
2	CAL2	1	Water	NA	NA	24.7	7.01	07/21/2020	16:37
3	CAL3	1	Water	NA	NA	24.7	10.02	07/21/2020	16:39
4	ICV	1	Water	NA	NA	24.7	7.00	07/21/2020	16:41
5	CCV1	1	Water	NA	NA	24.7	2.01	07/21/2020	16:43
6	L3362-05	1	Water	NA	NA	24.6	6.59	07/21/2020	16:46
7	L3362-05DUP	1	Water	NA	NA	24.6	6.57	07/21/2020	16:49
8	L3363-03	1	Water	NA	NA	25.5	7.75	07/21/2020	16:52
9	L3363-05	1	Water	NA	NA	25.1	5.52	07/21/2020	16:55
10	L3370-09	1	Water	NA	NA	23.7	6.14	07/21/2020	16:58
11	CCV2	1	Water	NA	NA	24.7	12.02	07/21/2020	17:00

LB110128

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Phwater-72120 WorkList ID : 140633 Department : Wet-Chemistry Date : 07-21-2020 11:16:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/21/2020	Water	L3362-05 F	pH	Cool 4 deg C	PSEG01	H31	1754	07/21/2020	9040C
07/21/2020	Water	L3363-03 A	pH	Cool 4 deg C	PSEG01	H31	79	07/21/2020	9040C
07/21/2020	Water	L3363-05 A	pH	Cool 4 deg C	PSEG01	H31	80	07/21/2020	9040C
07/21/2020	Water	L3370-09 A	pH	Cool 4 deg C	PSEG01	H31	TP-3-GW	07/21/2020	9040C

Date/Time

7/21/2020 16:30

Received by:

AMLwet-chem

Relinquished by:

KS (wet-chem)

Date/Time

7/21/2020 17:10

Received by:

KS (wet-chem)

Relinquished by:

AMLwet-chem