



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Alexander

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110133

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	24.7	4.01	07/21/2020	16:35
2	CAL2	1	Water	NA	NA	24.7	7.01	07/21/2020	16:37
3	CAL3	1	Water	NA	NA	24.7	10.02	07/21/2020	16:39
4	ICV	1	Water	NA	NA	24.7	7.00	07/21/2020	16:41
5	CCV1	1	Water	NA	NA	24.7	2.01	07/21/2020	16:43
6	L3370-10	1	Water	NA	NA	23.7	6.14	07/21/2020	16:58
7	L3370-10DUP	1	Water	NA	NA	23.7	6.15	07/21/2020	17:02
8	CCV2	1	Water	NA	NA	24.7	12.02	07/21/2020	17:05

26110133

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Corrosivitywater-72120 WorkList ID : 140688 Department : Wet-Chemistry Date : 07-21-2020 12:01:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/21/2020	Water	L3370-10 A	Corrosivity	Cool 4 deg C	PSEG01	H31	TP-3-GW	07/21/2020	9040C

Date/Time 7/21/2020 16:30
 Received by: AM (wet-chem)
 Relinquished by: KS (wet-chem)

Date/Time 7/21/2020 17:10
 Received by: KS (wet-chem)
 Relinquished by: AM (wet-chem)