


LB110130

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.0422e-004

C= 8.3078e-003

Rho= 0.9998013

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	43	0.000	43				
0.2	0.200	0.202	0.002	950	0.0 %	950				
2.5	2.500	2.435	-0.065	11881	0.0 %	11881				
5.0	5.000	4.995	-0.005	24420	0.0 %	24420				
7.5	7.500	7.641	0.141	37372	0.0 %	37372				
10.0	10.000	9.910	-0.090	48484	0.0 %	48484				

LB110138 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Jul 2020 13:07:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	43	-	0.0007	7471B	PPB	22 Jul 2020 13:42:04	S	Std
0.2 - 1	50.2	950	-	0.20007	7471B	PPB	22 Jul 2020 13:44:21	S	Std
2.5 - 1	52.5	11881	-	2.50007	7471B	PPB	22 Jul 2020 13:46:38	S	Std
5.0 - 1	55	24420	-	5.00007	7471B	PPB	22 Jul 2020 13:48:55	S	Std
7.5 - 1	57.5	37372	-	7.50007	7471B	PPB	22 Jul 2020 13:53:48	S	Std
10.0 - 1	510	48484	-	10.00007	7471B	PPB	22 Jul 2020 13:58:41	S	Std
ICV36 - 1	ICV36	20440	4.1826	-7471B	PPB	PPB	22 Jul 2020 14:01:23	U	SMPL
ICB36 - 1	ICB36	-291	-0.0511	-7471B	PPB	PPB	22 Jul 2020 14:03:39	U	SMPL
CCV39 - 1	CCV39	24623	5.0369	-7471B	PPB	PPB	22 Jul 2020 14:05:55	U	SMPL
CCB39 - 1	CCB39	-296	-0.0521	-7471B	PPB	PPB	22 Jul 2020 14:08:10	U	SMPL
CRA - 1	CRA	905	0.1931	-7471B	PPB	PPB	22 Jul 2020 14:10:27	U	SMPL
HighStd - 1	HighStd	48479	9.9089	-7471B	PPB	PPB	22 Jul 2020 14:12:42	U	SMPL
ChkStd - 1	ChkStd	35178	7.1925	-7471B	PPB	PPB	22 Jul 2020 14:14:57	U	SMPL
PB130375BL - 1	PBS	-271	-0.0470	-7471B	PPB	PPB	22 Jul 2020 14:17:18	U	SMPL
PB130375BS - 1	LCSS	18414	3.7689	-7471B	PPB	PPB	22 Jul 2020 14:19:36	U	SMPL
L3186-06 - 1	OR-03-072120	5934	1.2202	-7471B	PPB	PPB	22 Jul 2020 14:26:48	U	SMPL
L3186-06DUP - 1	OR-03-072120DUP	4874	1.0037	-7471B	PPB	PPB	22 Jul 2020 14:29:06	U	SMPL
L3186-06MS - 1	OR-03-072120MS	24333	4.9777	-7471B	PPB	PPB	22 Jul 2020 14:31:25	U	SMPL
L3186-06MSD - 1	OR-03-072120MSD	24616	5.0355	-7471B	PPB	PPB	22 Jul 2020 14:33:44	U	SMPL
L3187-09 - 1	SU-04-072120	861	0.1841	-7471B	PPB	PPB	22 Jul 2020 14:36:02	U	SMPL
CCV40 - 1	CCV40	24725	5.0577	-7471B	PPB	PPB	22 Jul 2020 14:38:21	U	SMPL
CCB40 - 1	CCB40	-72	-0.0064	-7471B	PPB	PPB	22 Jul 2020 14:40:37	U	SMPL
L3189-01 - 1	BU-02-072120	2170	0.4515	-7471B	PPB	PPB	22 Jul 2020 14:42:54	U	SMPL
L3360-01 - 1	SP-1A	4787	0.9859	-7471B	PPB	PPB	22 Jul 2020 14:45:09	U	SMPL
L3360-02 - 1	SP-1B	4060	0.8375	-7471B	PPB	PPB	22 Jul 2020 14:47:25	U	SMPL
L3360-03 - 1	SP-2	3548	0.7329	-7471B	PPB	PPB	22 Jul 2020 14:49:43	U	SMPL
L3360-04 - 1	SP-3	5897	1.2126	-7471B	PPB	PPB	22 Jul 2020 14:52:00	U	SMPL
L3362-02 - 1	SOIL-PILE	22251	4.5525	-7471B	PPB	PPB	22 Jul 2020 14:54:18	U	SMPL
L3362-04 - 1	RBR-59	9842	2.0183	-7471B	PPB	PPB	22 Jul 2020 14:56:35	U	SMPL
L3365-02 - 1	20017-20405	5188	1.0678	-7471B	PPB	PPB	22 Jul 2020 14:58:54	U	SMPL
L3365-04 - 1	52-09-57	40851	8.3511	-7471B	PPB	PPB	22 Jul 2020 15:01:13	U	SMPL
L3366-02 - 1	SP-1A	1839	0.3839	-7471B	PPB	PPB	22 Jul 2020 15:03:30	U	SMPL
CCV41 - 1	CCV41	25799	5.2771	-7471B	PPB	PPB	22 Jul 2020 15:05:48	U	SMPL
CCB41 - 1	CCB41	-87	-0.0095	-7471B	PPB	PPB	22 Jul 2020 15:08:05	U	SMPL
L3366-06 - 1	SP-2A	2942	0.6091	-7471B	PPB	PPB	22 Jul 2020 15:10:22	U	SMPL
L3368-01 - 1	20200435-COMPOSITE	6664	1.3693	-7471B	PPB	PPB	22 Jul 2020 15:12:37	U	SMPL
L3370-03 - 1	TP-3-S	14320	2.9328	-7471B	PPB	PPB	22 Jul 2020 15:14:54	U	SMPL
L3370-07 - 1	TP-3-D	7437	1.5271	-7471B	PPB	PPB	22 Jul 2020 15:17:12	U	SMPL
L3372-01 - 1	FK-BPM-04-011-ABCD	1360	0.2861	-7471B	PPB	PPB	22 Jul 2020 15:19:29	U	SMPL
L3373-01 - 1	FK-BPM-03-016-ABCD	1824	0.3808	-7471B	PPB	PPB	22 Jul 2020 15:21:46	U	SMPL
L3374-01 - 1	FK-BPM-01-C2-C3-C2A-C3A	1513	0.3173	-7471B	PPB	PPB	22 Jul 2020 15:24:02	U	SMPL
L3186-06LX5 - 1		870	0.1860	-7471B	PPB	PPB	22 Jul 2020 15:26:19	U	SMPL
L3186-06A - 1		25245	5.1639	-7471B	PPB	PPB	22 Jul 2020 15:31:32	U	SMPL
CCV42 - 1	CCV42	25845	5.2865	-7471B	PPB	PPB	22 Jul 2020 15:33:47	U	SMPL
CCB42 - 1	CCB42	-101	-0.0123	-7471B	PPB	PPB	22 Jul 2020 15:36:06	U	SMPL