



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110147

Slope : 99.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	07/22/2020	18:00
2	CAL2	1	Water	NA	NA	20.4	7.00	07/22/2020	18:01
3	CAL3	1	Water	NA	NA	20.4	10.03	07/22/2020	18:02
4	ICV	1	Water	NA	NA	20.4	7.01	07/22/2020	18:03
5	CCV1	1	Water	NA	NA	20.4	2.02	07/22/2020	18:05
6	L3378-17	1	Water	NA	NA	26.3	11.57	07/22/2020	18:06
7	L3378-17DUP	1	Water	NA	NA	26.4	11.58	07/22/2020	18:07
8	CCV2	1	Water	NA	NA	20.5	12.02	07/22/2020	18:10

20118148

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3378 WorkList ID : 140699 Department : Wet-Chemistry Date : 07-22-2020 16:53:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/22/2020	Water	L3378-17	Corrosivity	Cool 4 deg C	PSEG01	H43	DRUM-WATER	07/22/2020	9040C

Date/Time 07/22/2020 18:00
Received by: Jo (wet) (mem)
Relinquished by: Jo

Date/Time 07/22/2020 18:00
Received by: Jo
Relinquished by: Jo (wet) (mem)