



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110148

Slope : 99.4

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2629         |
| buffer solution pH 7 yellow    | W2626         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2632         |
| buffer solution pH 7 yellow    | W2569         |
| Buffer Solution, PH2 (500ml)   | W2592         |
| Buffer Solution, PH12 (500ml)  | W2616         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.4             | 4.01        | 07/22/2020 | 18:00     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.4             | 7.00        | 07/22/2020 | 18:01     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.4             | 10.04       | 07/22/2020 | 18:02     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.4             | 7.01        | 07/22/2020 | 18:03     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.4             | 2.02        | 07/22/2020 | 18:05     |
| 6   | L3378-16    | 1  | Water  | NA          | NA          | 26.3             | 11.57       | 07/22/2020 | 18:07     |
| 7   | L3378-16DUP | 1  | Water  | NA          | NA          | 26.4             | 11.58       | 07/22/2020 | 18:08     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.5             | 12.02       | 07/22/2020 | 18:10     |

20118148

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L3378      WorkList ID : 140699      Department : Wet-Chemistry      Date : 07-22-2020 16:53:00

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
| 07/22/2020 | Water  | L3378-17 | Corrosivity | Cool 4 deg C | PSEG01   | H43              | DRUM-WATER      | 07/22/2020   | 9040C  |

Date/Time 07/22/2020 18:00  
Received by: J. Wet  
Relinquished by: J. Wet

Date/Time 07/22/2020 18:00  
Received by: J. Wet  
Relinquished by: J. Wet