

LB110161


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36938



A= 0.0000e+000

B= 2.7191e-004

C= -2.9760e-003

Rho= 0.9999452

Accept = Accepted

slope
 $y = mx + c$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.016	0.016	68	0.000	68					1
0.05	0.050					289					15
0.20	0.200	0.230	0.030	855	0.0 %	855					0
2.50	2.500	2.495	-0.005	9187	0.0 %	9187					0
5.00	5.000	4.922	-0.078	18111	0.0 %	18111					-2
7.50	7.500	7.497	-0.003	27583	0.0 %	27583					0
10.0	10.000	10.041	0.041	36938	0.0 %	36938					0

LB110153 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Jul 2020 09:46:09

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-110	-	0.00007470A	PPB	23 Jul 2020 10:16:33	S	Std
0.2 - 1	50.2	611	-	0.20007470A	PPB	23 Jul 2020 10:18:40	S	Std
2.5 - 1	52.5	7838	-	2.50007470A	PPB	23 Jul 2020 10:20:48	S	Std
5.0 - 1	15	16451	-	5.00007470A	PPB	23 Jul 2020 10:22:56	S	Std
7.5 - 1	57.5	23745	-	7.50007470A	PPB	23 Jul 2020 10:25:03	S	Std
10.0 - 1	51.0	32139	-	10.00007470A	PPB	23 Jul 2020 10:27:10	S	Std
ICV37 - 1	ICV37	12565	3.9261	-7470A	PPB	23 Jul 2020 10:29:51	U	SMPL
ICB37 - 1	ICB37	-349	-0.0899	-7470A	PPB	23 Jul 2020 10:31:56	U	SMPL
CCV43 - 1	CCV43	16577	5.1738	-7470A	PPB	23 Jul 2020 10:34:03	U	SMPL
CCB43 - 1	CCB43	-368	-0.0958	-7470A	PPB	23 Jul 2020 10:36:08	U	SMPL
CRA - 1	CRA	533	0.1844	-7470A	PPB	23 Jul 2020 10:38:16	U	SMPL
HighStd - 1	HighStd	32207	10.0345	-7470A	PPB	23 Jul 2020 10:40:22	U	SMPL
ChkStd - 1	ChkStd	23205	7.2350	-7470A	PPB	23 Jul 2020 10:42:28	U	SMPL
PB130388BL - 1	PBW	-141	-0.0252	-7470A	PPB	23 Jul 2020 10:47:00	U	SMPL
PB130388BS - 1	LCSW	11859	3.7066	-7470A	PPB	23 Jul 2020 10:49:07	U	SMPL
L3186-10 - 1	OR-03-072120	-295	-0.0731	-7470A	PPB	23 Jul 2020 10:51:14	U	SMPL
L3186-10DUP - 1	OR-03-072120DUP	-150	-0.0280	-7470A	PPB	23 Jul 2020 10:53:22	U	SMPL
L3186-10MS - 1	OR-03-072120MS	12196	3.8114	-7470A	PPB	23 Jul 2020 10:55:30	U	SMPL
L3186-10MSD - 1	OR-03-072120MSD	11618	3.6316	-7470A	PPB	23 Jul 2020 10:57:37	U	SMPL
L3187-13 - 1	SU-04-072120	-338	-0.0865	-7470A	PPB	23 Jul 2020 10:59:43	U	SMPL
CCV44 - 1	CCV44	16667	5.2018	-7470A	PPB	23 Jul 2020 11:01:49	U	SMPL
CCB44 - 1	CCB44	-387	-0.1017	-7470A	PPB	23 Jul 2020 11:03:54	U	SMPL
L3189-04 - 1	BU-02-072120	-155	-0.0296	-7470A	PPB	23 Jul 2020 11:06:02	U	SMPL
L3363-02 - 1	81-83	-109	-0.0153	-7470A	PPB	23 Jul 2020 11:08:07	U	SMPL
L3366-12 - 1	SP-1A	-97	-0.0115	-7470A	PPB	23 Jul 2020 11:10:13	U	SMPL
L3366-13 - 1	SP-2A	-77	-0.0053	-7470A	PPB	23 Jul 2020 11:12:20	U	SMPL
L3370-04 - 1	TP-3-S	-72	-0.0038	-7470A	PPB	23 Jul 2020 11:14:26	U	SMPL
L3370-08 - 1	TP-3-D	-60	-0.0000	-7470A	PPB	23 Jul 2020 11:16:34	U	SMPL
PB130389BL - 1	PBW	-48	0.0037	-7470A	PPB	23 Jul 2020 11:18:41	U	SMPL
PB130389BS - 1	LCSW	11624	3.6335	-7470A	PPB	23 Jul 2020 11:20:48	U	SMPL
L3362-05 - 1	1754	-140	-0.0249	-7470A	PPB	23 Jul 2020 11:22:56	U	SMPL
L3363-03 - 1	79	-72	-0.0038	-7470A	PPB	23 Jul 2020 11:25:02	U	SMPL
CCV45 - 1	CCV45	16814	5.2475	-7470A	PPB	23 Jul 2020 11:27:09	U	SMPL
CCB45 - 1	CCB45	-363	-0.0943	-7470A	PPB	23 Jul 2020 11:29:15	U	SMPL
L3370-10 - 1	TP-3-GW	-147	-0.0271	-7470A	PPB	23 Jul 2020 11:31:22	U	SMPL
L3370-10DUP - 1	TP-3-GWDUP	-91	-0.0097	-7470A	PPB	23 Jul 2020 11:33:29	U	SMPL
L3370-10MS - 1	TP-3-GWMS	12636	3.9482	-7470A	PPB	23 Jul 2020 11:35:35	U	SMPL
L3370-10MSD - 1	TP-3-GWMSD	12411	3.8782	-7470A	PPB	23 Jul 2020 11:37:41	U	SMPL
PB130390BL - 1	PBW	-324	-0.0821	-7470A	PPB	23 Jul 2020 11:39:48	U	SMPL
PB130390BS - 1	LCSW	13004	4.0626	-7470A	PPB	23 Jul 2020 11:44:36	U	SMPL
L3370-09 - 1	TP-3-GW	1432	0.4640	-7470A	PPB	23 Jul 2020 11:49:14	U	SMPL
L3370-09DUP - 1	TP-3-GWDUP	1217	0.3971	-7470A	PPB	23 Jul 2020 11:51:21	U	SMPL
L3370-09MS - 1	TP-3-GWMS	16034	5.0049	-7470A	PPB	23 Jul 2020 11:53:29	U	SMPL
L3370-09MSD - 1	TP-3-GWMSD	14861	4.6401	-7470A	PPB	23 Jul 2020 12:00:50	U	SMPL
CCV46 - 1	CCV46	16033	5.0046	-7470A	PPB	23 Jul 2020 12:02:56	U	SMPL
CCB46 - 1	CCB46	-319	-0.0806	-7470A	PPB	23 Jul 2020 12:05:04	U	SMPL
PB130388TB - 1	PB130361TB	-137	-0.0240	-7470A	PPB	23 Jul 2020 12:07:11	U	SMPL
PB130389TB - 1	PB130363TB	-69	-0.0028	-7470A	PPB	23 Jul 2020 12:09:17	U	SMPL
L3186-10LX5 - 1		-76	-0.0050	-7470A	PPB	23 Jul 2020 12:11:24	U	SMPL
L3186-10A - 1		13228	4.1323	-7470A	PPB	23 Jul 2020 12:13:31	U	SMPL
L3370-10LX5 - 1		-271	-0.0657	-7470A	PPB	23 Jul 2020 12:15:37	U	SMPL
L3370-10A - 1		12311	3.8471	-7470A	PPB	23 Jul 2020 12:17:44	U	SMPL
L3370-09LX5 - 1		-228	-0.0523	-7470A	PPB	23 Jul 2020 12:19:51	U	SMPL
L3370-09A - 1		14459	4.5151	-7470A	PPB	23 Jul 2020 12:21:59	U	SMPL
CCV47 - 1	CCV47	16163	5.0450	-7470A	PPB	23 Jul 2020 12:24:07	U	SMPL
CCB47 - 1	CCB47	-399	-0.1055	-7470A	PPB	23 Jul 2020 12:26:15	U	SMPL