



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110174

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/23/2020	16:40
2	CAL2	1	Water	NA	NA	20.3	7.00	07/23/2020	16:41
3	CAL3	1	Water	NA	NA	20.3	10.03	07/23/2020	16:42
4	ICV	1	Water	NA	NA	20.3	6.98	07/23/2020	16:44
5	CCV1	1	Water	NA	NA	20.3	2.01	07/23/2020	16:45
6	L3382-03	1	Solid	20.03	20	24.5	8.53	07/23/2020	16:46
7	L3382-07	1	Solid	20.04	20	24.6	5.69	07/23/2020	16:47
8	L3382-11	1	Solid	20.02	20	24.4	8.73	07/23/2020	16:48
9	L3382-15	1	Solid	20.03	20	24.3	6.22	07/23/2020	16:50
10	L3382-19	1	Solid	20.04	20	24.4	7.30	07/23/2020	17:00
11	L3382-23	1	Solid	20.02	20	24.5	7.70	07/23/2020	17:01
12	L3383-03	1	Solid	20.03	20	24.4	6.56	07/23/2020	17:03
13	L3383-07	1	Solid	20.04	20	24.3	7.36	07/23/2020	17:05
14	L3383-11	1	Solid	20.02	20	24.4	5.73	07/23/2020	17:06
15	L3383-15	1	Solid	20.03	20	24.2	9.19	07/23/2020	17:10
16	CCV2	1	Water	NA	NA	20.3	12.02	07/23/2020	17:11
17	L3383-19	1	Solid	20.05	20	23.9	7.30	07/23/2020	17:12
18	L3383-23	1	Solid	20.02	20	24.1	8.93	07/23/2020	17:15
19	L3383-27	1	Solid	20.03	20	24.3	7.91	07/23/2020	17:19
20	L3383-31	1	Solid	20.04	20	24.1	7.35	07/23/2020	17:20

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	L3414-04	1	Solid	20.03	20	25.6	8.80	07/23/2020	17:22
22	L3414-04DUP	1	Solid	20.04	20	25.7	8.81	07/23/2020	17:23
23	CCV3	1	Water	NA	NA	20.3	2.01	07/23/2020	17:25

28110174

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3382 WorkList ID : 140740 Department : Wet-Chemistry Date : 07-23-2020 11:43:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/22/2020	Solid	L3382-03	Corrosivity	Cool 4 deg C	PSEG01	H51	CTR-007	07/22/2020	9045D
07/22/2020	Solid	L3382-07	Corrosivity	Cool 4 deg C	PSEG01	H51	CTR-020	07/22/2020	9045D
07/22/2020	Solid	L3382-11	Corrosivity	Cool 4 deg C	PSEG01	H51	CTR-023	07/22/2020	9045D
07/22/2020	Solid	L3382-15	Corrosivity	Cool 4 deg C	PSEG01	H51	CTR-005	07/22/2020	9045D
07/22/2020	Solid	L3382-19	Corrosivity	Cool 4 deg C	PSEG01	H51	CTR-011	07/22/2020	9045D
07/22/2020	Solid	L3382-23	Corrosivity	Cool 4 deg C	PSEG01	H51	CTR-024	07/22/2020	9045D
07/22/2020	Solid	L3383-03	Corrosivity	Cool 4 deg C	PSEG01	I11	CRT-013	07/22/2020	9045D
07/22/2020	Solid	L3383-07	Corrosivity	Cool 4 deg C	PSEG01	I11	CRT-018	07/22/2020	9045D
07/22/2020	Solid	L3383-11	Corrosivity	Cool 4 deg C	PSEG01	I11	CRT-009	07/22/2020	9045D
07/22/2020	Solid	L3383-15	Corrosivity	Cool 4 deg C	PSEG01	I11	CRT-022	07/22/2020	9045D
07/22/2020	Solid	L3383-19	Corrosivity	Cool 4 deg C	PSEG01	I11	CRT-008	07/22/2020	9045D
07/22/2020	Solid	L3383-23	Corrosivity	Cool 4 deg C	PSEG01	I11	CRT-015	07/22/2020	9045D
07/22/2020	Solid	L3383-27	Corrosivity	Cool 4 deg C	PSEG01	I11	CRT-014	07/22/2020	9045D
07/22/2020	Solid	L3383-31	Corrosivity	Cool 4 deg C	PSEG01	I11	CRT-002	07/22/2020	9045D
07/23/2020	Solid	L3414-04	Corrosivity	Cool 4 deg C	PSEG01	I13	61820	07/23/2020	9045D

Date/Time 07/23/2020 18:00
 Received by: KS (webchem)
 Relinquished by: JO (webchem)

Date/Time 07/23/2020 13:40
 Received by: JO (webchem)
 Relinquished by: KS (webchem)